

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061424\
 Data File : VW029160.D
 Acq On : 14 Jun 2024 17:24
 Operator : SY/MD
 Sample : P2871-11
 Misc : 5.19g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CJ1

Quant Time: Jun 15 00:23:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jun 15 00:20:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	308585	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	207558	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	50807	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	105928	21.608	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.440%
7) Chloroethane-d5	2.899	69	97530	25.908	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.640%
11) 1,1-Dichloroethene-d2	4.021	65	44622	20.709	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.840%
21) 2-Butanone-d5	7.075	46	54422	54.334	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.660%
24) Chloroform-d	7.648	84	227988	25.897	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	8.307	65	140854	30.390	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	121.560%
32) Benzene-d6	8.270	84	404936	31.786	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	127.160%
36) 1,2-Dichloropropane-d6	9.270	67	138571	35.876	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	143.520%#
41) Toluene-d8	10.325	98	290304	25.536	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.160%
43) trans-1,3-Dichloroprop...	10.575	79	33788	22.765	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.040%
47) 2-Hexanone-d5	10.922	63	33973	66.651	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	133.300%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	86833	32.360	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	129.440%#
66) 1,2-Dichlorobenzene-d4	13.848	152	44105	24.463	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.840%
Target Compounds						
16) Methylene chloride	4.917	84	17012	3.673	ug/L	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061424\
Data File : VW029160.D
Acq On : 14 Jun 2024 17:24
Operator : SY/MD
Sample : P2871-11
Misc : 5.19g/10mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CJ1

Quant Time: Jun 15 00:23:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jun 15 00:20:11 2024
Response via : Initial Calibration

