

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061424\
 Data File : VW029148.D
 Acq On : 14 Jun 2024 10:15
 Operator : SY/MD
 Sample : P2844-09
 Misc : 4.37g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4C61

Quant Time: Jun 15 00:20:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jun 15 00:20:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	404075	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	317548	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	98941	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	127296	19.830	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.320%
7) Chloroethane-d5	2.899	69	111387	22.597	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.400%
11) 1,1-Dichloroethene-d2	4.021	65	55454	19.655	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.600%
21) 2-Butanone-d5	7.075	46	83594	63.737	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	127.480%
24) Chloroform-d	7.648	84	265783	23.056	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.240%
26) 1,2-Dichloroethane-d4	8.307	65	161466	26.605	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.400%
32) Benzene-d6	8.270	84	497949	25.549	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.200%
36) 1,2-Dichloropropane-d6	9.270	67	164565	27.848	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	111.400%
41) Toluene-d8	10.318	98	399986	22.997	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.000%
43) trans-1,3-Dichloroprop...	10.575	79	57380	25.269	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	101.080%
47) 2-Hexanone-d5	10.922	63	50343	64.557	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	129.120%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	115028	28.019	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.080%
66) 1,2-Dichlorobenzene-d4	13.848	152	85835	24.447	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061424\
Data File : VW029148.D
Acq On : 14 Jun 2024 10:15
Operator : SY/MD
Sample : P2844-09
Misc : 4.37g/10mL/MSVOA_W/SOIL/B
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4C61

Quant Time: Jun 15 00:20:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jun 15 00:20:11 2024
Response via : Initial Calibration

