

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061824\
 Data File : VW029194.D
 Acq On : 18 Jun 2024 21:37
 Operator : SY/MD
 Sample : P2934-10
 Misc : 5.14g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CP6

Quant Time: Jun 19 01:53:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 19 01:46:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	325262	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	255751	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	73358	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	94914	21.083	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.320%
7) Chloroethane-d5	2.899	69	85917	22.599	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.400%
11) 1,1-Dichloroethene-d2	4.021	65	42518	19.560	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.240%
21) 2-Butanone-d5	7.081	46	70108	52.050	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.100%
24) Chloroform-d	7.648	84	211937	21.843	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.360%
26) 1,2-Dichloroethane-d4	8.307	65	126182	22.651	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.600%
32) Benzene-d6	8.276	84	394694	25.408	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.640%
36) 1,2-Dichloropropane-d6	9.270	67	132815	26.896	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	107.600%
41) Toluene-d8	10.324	98	304602	22.163	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.640%
43) trans-1,3-Dichloroprop...	10.574	79	39385	19.792	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.160%
47) 2-Hexanone-d5	10.922	63	46437	59.989	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.980%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	97193	24.730	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.920%
66) 1,2-Dichlorobenzene-d4	13.848	152	63327	24.870	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.480%
Target Compounds						
					Qvalue	
13) Acetone	4.118	43	25013	21.852	ug/L	97
14) Carbon disulfide	4.393	76	9241	0.751	ug/L	99
16) Methylene chloride	4.923	84	17933	3.136	ug/L	99
22) 2-Butanone	7.179	43	9978	6.359	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061824\
Data File : VW029194.D
Acq On : 18 Jun 2024 21:37
Operator : SY/MD
Sample : P2934-10
Misc : 5.14g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CP6

Quant Time: Jun 19 01:53:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 19 01:46:42 2024
Response via : Initial Calibration

