

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW061923\  
 Data File : VW026254.D  
 Acq On : 19 Jun 2023 17:56  
 Operator : SY/MD  
 Sample : 03325-02MS  
 Misc : 5.28g/10mL/MSVOA\_W/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 YBW38MS

Manual Integrations  
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/20/2023  
 Supervised By : Mahesh Dadoda 06/26/2023

Quant Time: Jun 20 03:19:03 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM053023SMA.M  
 Quant Title : SFAM01.0  
 QLast Update : Fri Jun 16 23:58:43 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min)  |
|-------------------------------|--------|-------|----------|----------|--------|-----------|
| -----                         |        |       |          |          |        |           |
| Internal Standards            |        |       |          |          |        |           |
| 1) 1,4-Difluorobenzene        | 8.843  | 114   | 738868   | 25.000   | ug/L   | 0.00      |
| 28) Chlorobenzene-d5          | 11.629 | 117   | 611313   | 25.000   | ug/L   | 0.00      |
| 58) 1,4-Dichlorobenzene-d4    | 13.555 | 152   | 222678   | 25.000   | ug/L   | 0.00      |
| System Monitoring Compounds   |        |       |          |          |        |           |
| 4) Vinyl Chloride-d3          | 2.362  | 65    | 180167   | 17.333   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 30 - 150 | Recovery | =      | 69.320%   |
| 7) Chloroethane-d5            | 2.893  | 69    | 178055   | 22.883   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 30 - 150 | Recovery | =      | 91.520%   |
| 11) 1,1-Dichloroethene-d2     | 4.021  | 65    | 94059    | 18.352   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 45 - 110 | Recovery | =      | 73.400%   |
| 21) 2-Butanone-d5             | 7.075  | 46    | 216667   | 71.598   | ug/L   | 0.00      |
| Spiked Amount                 | 50.000 | Range | 20 - 135 | Recovery | =      | 143.200%# |
| 24) Chloroform-d              | 7.648  | 84    | 514172   | 23.764   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 40 - 150 | Recovery | =      | 95.040%   |
| 26) 1,2-Dichloroethane-d4     | 8.306  | 65    | 282884   | 23.910   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 70 - 130 | Recovery | =      | 95.640%   |
| 32) Benzene-d6                | 8.276  | 84    | 961584   | 23.377   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 20 - 135 | Recovery | =      | 93.520%   |
| 36) 1,2-Dichloropropane-d6    | 9.270  | 67    | 327387   | 25.122   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 70 - 120 | Recovery | =      | 100.480%  |
| 41) Toluene-d8                | 10.324 | 98    | 782299   | 21.445   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 30 - 130 | Recovery | =      | 85.760%   |
| 43) trans-1,3-Dichloroprop... | 10.574 | 79    | 99694    | 19.305   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 30 - 135 | Recovery | =      | 77.240%   |
| 47) 2-Hexanone-d5             | 10.922 | 63    | 141536   | 68.879   | ug/L   | 0.00      |
| Spiked Amount                 | 50.000 | Range | 20 - 135 | Recovery | =      | 137.760%# |
| 56) 1,1,2,2-Tetrachloroeth... | 12.690 | 84    | 277949   | 29.310   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 45 - 120 | Recovery | =      | 117.240%  |
| 66) 1,2-Dichlorobenzene-d4    | 13.848 | 152   | 185152   | 21.723   | ug/L   | 0.00      |
| Spiked Amount                 | 25.000 | Range | 75 - 120 | Recovery | =      | 86.880%   |
| Target Compounds              |        |       |          |          |        |           |
|                               |        |       |          |          | Qvalue |           |
| 2) Dichlorodifluoromethane    | 2.015  | 85    | 201771m  | 22.276   | ug/L   |           |
| 3) Chloromethane              | 2.222  | 50    | 248199   | 21.371   | ug/L   | 97        |
| 5) Vinyl chloride             | 2.375  | 62    | 253990   | 20.460   | ug/L   | 99        |
| 6) Bromomethane               | 2.783  | 94    | 142179   | 21.009   | ug/L   | 100       |
| 8) Chloroethane               | 2.929  | 64    | 158065   | 23.562   | ug/L   | 96        |
| 9) Trichlorofluoromethane     | 2.359  | 101   | 195885   | 24.527   | ug/L   | 99        |
| 10) 1,1,2-Trichloro-1,2,2-... | 4.069  | 101   | 214848   | 21.319   | ug/L   | 98        |
| 12) 1,1-Dichloroethene        | 4.045  | 96    | 192353   | 20.447   | ug/L   | 89        |
| 13) Acetone                   | 4.118  | 43    | 155915   | 66.586   | ug/L   | 97        |
| 14) Carbon disulfide          | 4.386  | 76    | 490339   | 14.765   | ug/L   | 99        |
| 15) Methyl Acetate            | 4.667  | 43    | 140365   | 26.709   | ug/L   | 99        |
| 16) Methylene chloride        | 4.917  | 84    | 287527   | 24.381   | ug/L   | 98        |
| 17) trans-1,2-Dichloroethene  | 5.423  | 96    | 189751   | 17.978   | ug/L   | 98        |
| 18) Methyl tert-butyl Ether   | 5.423  | 73    | 379142   | 22.792   | ug/L   | 100       |
| 19) 1,1-Dichloroethane        | 6.215  | 63    | 480853   | 22.378   | ug/L   | 99        |
| 20) cis-1,2-Dichloroethene    | 7.166  | 96    | 229265   | 20.506   | ug/L   | 97        |
| 22) 2-Butanone                | 7.166  | 43    | 214902   | 64.338   | ug/L   | 97        |
| 23) Bromochloromethane        | 7.514  | 128   | 101537   | 21.295   | ug/L   | 96        |
| 25) Chloroform                | 7.672  | 83    | 450663   | 22.517   | ug/L   | 100       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW061923\  
 Data File : VW026254.D  
 Acq On : 19 Jun 2023 17:56  
 Operator : SY/MD  
 Sample : 03325-02MS  
 Misc : 5.28g/10mL/MSVOA\_W/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 YBW38MS

Manual Integrations  
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/20/2023  
 Supervised By :Mahesh Dadoda 06/26/2023

Quant Time: Jun 20 03:19:03 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM053023SMA.M  
 Quant Title : SFAM01.0  
 QLast Update : Fri Jun 16 23:58:43 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 27) 1,2-Dichloroethane        | 8.398  | 62   | 304983   | 22.095 | ug/L  | 96       |
| 29) Cyclohexane               | 7.959  | 56   | 392958   | 20.821 | ug/L  | 100      |
| 30) 1,1,1-Trichloroethane     | 7.874  | 97   | 349758   | 23.577 | ug/L  | 100      |
| 31) Carbon tetrachloride      | 8.069  | 117  | 279622   | 20.655 | ug/L  | 99       |
| 33) Benzene                   | 8.319  | 78   | 952840   | 22.753 | ug/L  | 100      |
| 34) Trichloroethene           | 9.093  | 95   | 220937   | 20.604 | ug/L  | 95       |
| 35) Methylcyclohexane         | 9.337  | 83   | 362550   | 19.107 | ug/L  | 97       |
| 37) 1,2-Dichloropropane       | 9.367  | 63   | 270624   | 23.859 | ug/L  | 99       |
| 38) Bromodichloromethane      | 9.642  | 83   | 301232   | 22.859 | ug/L  | 98       |
| 39) cis-1,3-Dichloropropene   | 10.068 | 75   | 346217   | 20.771 | ug/L  | 99       |
| 40) 4-Methyl-2-pentanone      | 10.209 | 43   | 402149   | 60.137 | ug/L  | 99       |
| 42) Toluene                   | 10.385 | 91   | 924043   | 21.573 | ug/L  | 97       |
| 44) trans-1,3-Dichloropropene | 10.605 | 75   | 275951   | 19.872 | ug/L  | 98       |
| 45) 1,1,2-Trichloroethane     | 10.788 | 97   | 179879   | 24.063 | ug/L  | 98       |
| 46) Tetrachloroethene         | 10.861 | 164  | 144362   | 19.033 | ug/L  | 96       |
| 48) 2-Hexanone                | 10.964 | 43   | 289248   | 61.345 | ug/L  | 98       |
| 49) Dibromochloromethane      | 11.129 | 129  | 178186   | 22.735 | ug/L  | 92       |
| 50) 1,2-Dibromoethane         | 11.233 | 107  | 151526   | 21.891 | ug/L  | 97       |
| 51) Chlorobenzene             | 11.653 | 112  | 514253   | 19.160 | ug/L  | 99       |
| 52) Ethylbenzene              | 11.727 | 91   | 978037   | 20.242 | ug/L  | 99       |
| 53) m,p-Xylene                | 11.836 | 106  | 353301   | 19.655 | ug/L  | 92       |
| 54) o-Xylene                  | 12.159 | 106  | 349402   | 20.595 | ug/L  | 96       |
| 55) Styrene                   | 12.178 | 104  | 533610   | 18.334 | ug/L  | 96       |
| 57) 1,1,2,2-Tetrachloroethane | 12.714 | 83   | 226835   | 25.073 | ug/L  | 97       |
| 59) Bromoform                 | 12.348 | 173  | 95791    | 28.641 | ug/L  | 98       |
| 60) Isopropylbenzene          | 12.464 | 105  | 943192   | 26.963 | ug/L  | 99       |
| 61) 1,2,3-Trichloropropane    | 12.769 | 75   | 166395   | 33.360 | ug/L  | 99       |
| 62) 1,3,5-Trimethylbenzene    | 12.940 | 105  | 713668   | 25.664 | ug/L  | 98       |
| 63) 1,2,4-Trimethylbenzene    | 13.245 | 105  | 617637   | 24.250 | ug/L  | 99       |
| 64) 1,3-Dichlorobenzene       | 13.495 | 146  | 282041   | 19.176 | ug/L  | 95       |
| 65) 1,4-Dichlorobenzene       | 13.574 | 146  | 283847   | 19.169 | ug/L  | 98       |
| 67) 1,2-Dichlorobenzene       | 13.866 | 146  | 274177   | 20.759 | ug/L  | 99       |
| 68) 1,2-Dibromo-3-chloropr... | 14.482 | 75   | 33506    | 28.994 | ug/L  | 88       |
| 69) 1,3,5-Trichlorobenzene    | 14.622 | 180  | 143657   | 14.219 | ug/L  | 98       |
| 70) 1,2,4-trichlorobenzene    | 15.128 | 180  | 87044    | 10.441 | ug/L  | 98       |
| 71) Naphthalene               | 15.360 | 128  | 222813   | 14.995 | ug/L  | 99       |
| 72) 1,2,3-Trichlorobenzene    | 15.549 | 180  | 86964    | 11.936 | ug/L  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW061923\  
Data File : VW026254.D  
Acq On : 19 Jun 2023 17:56  
Operator : SY/MD  
Sample : 03325-02MS  
Misc : 5.28g/10mL/MSVOA\_W/SOIL  
ALS Vial : 1 Sample Multiplier: 1


Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS

Quant Time: Jun 20 03:19:03 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM053023SMA.M  
Quant Title : SFAM01.0  
QLast Update : Fri Jun 16 23:58:43 2023  
Response via : Initial Calibration

Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023





#2  
Dichlorodifluoromethane  
Concen: 22.276 ug/L m  
RT: 2.015 min Scan# 21  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion: 85 Resp: 20177  
Ion Ratio Lower Upper  
85 100  
87 22.6 22.2 33.2

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#3  
Chloromethane  
Concen: 21.371 ug/L  
RT: 2.222 min Scan# 55  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Tgt Ion: 50 Resp: 248199  
Ion Ratio Lower Upper  
50 100  
52 32.2 21.3 39.6





#5  
 Vinyl chloride  
 Concen: 20.460 ug/L  
 RT: 2.375 min Scan# 80  
 Delta R.T. -0.000 min  
 Lab File: VW026254.D  
 Acq: 19 Jun 2023 17:56




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 YBW38MS




Tgt Ion: 62 Resp: 253990  
 Ion Ratio Lower Upper  
 62 100  
 64 33.0 22.7 42.1

Manual Integrations  
 APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
 Supervised By :Mahesh Dadoda 06/26/2023




#6  
 Bromomethane  
 Concen: 21.009 ug/L  
 RT: 2.783 min Scan# 147  
 Delta R.T. -0.006 min  
 Lab File: VW026254.D  
 Acq: 19 Jun 2023 17:56




Tgt Ion: 94 Resp: 142179  
 Ion Ratio Lower Upper  
 94 100  
 96 91.6 64.2 119.2





#8  
Chloroethane  
Concen: 23.562 ug/L  
RT: 2.929 min Scan# 171  
Delta R.T. -0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS



Tgt Ion: 64 Resp: 15806  
Ion Ratio Lower Upper  
64 100  
66 31.1 23.2 43.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#9  
Trichlorofluoromethane  
Concen: 24.527 ug/L  
RT: 3.259 min Scan# 225  
Delta R.T. -0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Tgt Ion:101 Resp: 195885  
Ion Ratio Lower Upper  
101 100  
103 65.1 51.6 77.4









#10  
1,1,2-Trichloro-1,2,2-trifluoroethane  
Concen: 21.319 ug/L  
RT: 4.069 min Scan# 358  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS

Tgt Ion: 101 Resp: 21484  
Ion Ratio Lower Upper  
101 100  
85 43.9 36.1 54.1  
151 67.2 55.1 82.7




Manual Integrations  
APPROVED


Reviewed By : Semsettin Yesilyurt 06/20/2023  
Supervised By : Mahesh Dadoda 06/26/2023



#12  
1,1-Dichloroethene  
Concen: 20.447 ug/L  
RT: 4.045 min Scan# 354  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion: 96 Resp: 192353  
Ion Ratio Lower Upper  
96 100  
61 198.6 140.6 261.0  
63 158.4 131.7 244.5





#13  
Acetone  
Concen: 66.586 ug/L  
RT: 4.118 min Scan# 366  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS



Tgt Ion: 43 Resp: 155915  
Ion Ratio Lower Upper  
43 100  
58 30.6 0.0 58.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#14  
Carbon disulfide  
Concen: 14.765 ug/L  
RT: 4.386 min Scan# 410  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion: 76 Resp: 490339  
Ion Ratio Lower Upper  
76 100  
78 8.8 7.2 10.8





#15  
Methyl Acetate  
Concen: 26.709 ug/L  
RT: 4.667 min Scan# 456  
Delta R.T. -0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion: 43 Resp: 14036  
Ion Ratio Lower Upper  
43 100  
74 21.4 16.6 25.0

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#16  
Methylene chloride  
Concen: 24.381 ug/L  
RT: 4.917 min Scan# 497  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56



Tgt Ion: 84 Resp: 287527  
Ion Ratio Lower Upper  
84 100  
86 64.4 44.2 82.0  
49 143.3 98.8 183.6









#17  
trans-1,2-Dichloroethene  
Concen: 17.978 ug/L  
RT: 5.423 min Scan# 580  
Delta R.T. -0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS

Tgt Ion: 96 Resp: 18975  
Ion Ratio Lower Upper  
96 100  
61 159.4 113.0 209.8  
98 64.2 42.5 78.9




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#18  
Methyl tert-butyl Ether  
Concen: 22.792 ug/L  
RT: 5.423 min Scan# 580  
Delta R.T. -0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Tgt Ion: 73 Resp: 379142  
Ion Ratio Lower Upper  
73 100  
43 22.4 17.8 26.6  
57 23.7 18.9 28.3





#19

1,1-Dichloroethane

Concen: 22.378 ug/L

RT: 6.215 min Scan# 710

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56

Instrument :

MSVOA\_W

ClientSampleId :

YBW38MS

Tgt Ion: 63 Resp: 48085

Ion Ratio Lower Upper

63 100

65 31.5 22.3 41.5





83 12.9 9.1 16.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/20/2023

Supervised By :Mahesh Dadoda 06/26/2023



#20

cis-1,2-Dichloroethene

Concen: 20.506 ug/L

RT: 7.166 min Scan# 866

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56




Tgt Ion: 96 Resp: 229265


Ion Ratio Lower Upper

96 100

61 146.4 98.4 182.8


98 64.9 45.2 84.0





#22  
2-Butanone  
Concen: 64.338 ug/L  
RT: 7.166 min Scan# 866  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion: 43 Resp: 21490  
Ion Ratio Lower Upper  
43 100  
72 23.6 12.4 37.4

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#23  
Bromochloromethane  
Concen: 21.295 ug/L  
RT: 7.514 min Scan# 923  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Tgt Ion:128 Resp: 101537  
Ion Ratio Lower Upper  
128 100  
49 209.2 140.6 261.0  
130 122.8 86.0 159.6  
51 63.4 53.6 80.4





#25  
Chloroform  
Concen: 22.517 ug/L  
RT: 7.672 min Scan# 949  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS



Tgt Ion: 83 Resp: 45066  
Ion Ratio Lower Upper  
83 100  
85 64.1 45.1 83.7




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#27  
1,2-Dichloroethane  
Concen: 22.095 ug/L  
RT: 8.398 min Scan# 1068  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion: 62 Resp: 304983  
Ion Ratio Lower Upper  
62 100  
98 8.0 7.4 11.2





#29  
Cyclohexane  
Concen: 20.821 ug/L  
RT: 7.959 min Scan# 996  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion: 56 Resp: 392958  
Ion Ratio Lower Upper  
56 100  
69 28.8 23.3 34.9  
84 78.9 63.0 94.6

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#30  
1,1,1-Trichloroethane  
Concen: 23.577 ug/L  
RT: 7.874 min Scan# 982  
Delta R.T. 0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Tgt Ion: 97 Resp: 349758  
Ion Ratio Lower Upper  
97 100  
99 63.4 50.6 76.0  
61 51.0 40.5 60.7





#31  
Carbon tetrachloride  
Concen: 20.655 ug/L  
RT: 8.069 min Scan# 1014  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS



Tgt Ion:117 Resp: 27962  
Ion Ratio Lower Upper  
117 100  
119 95.8 77.4 116.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#33  
Benzene  
Concen: 22.753 ug/L  
RT: 8.319 min Scan# 1055  
Delta R.T. -0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion: 78 Resp: 952840





#34  
Trichloroethene  
Concen: 20.604 ug/L  
RT: 9.093 min Scan# 1182  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS



Tgt Ion: 95 Resp: 220937  
Ion Ratio Lower Upper  
95 100  
97 63.2 45.9 85.3  
132 94.5 67.1 124.7  
130 88.7 68.0 126.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#35  
Methylcyclohexane  
Concen: 19.107 ug/L  
RT: 9.337 min Scan# 1222  
Delta R.T. 0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Tgt Ion: 83 Resp: 362550  
Ion Ratio Lower Upper  
83 100  
55 88.4 68.2 102.4  
98 48.4 38.1 57.1





#37

1,2-Dichloropropane

Concen: 23.859 ug/L

RT: 9.367 min Scan# 112

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56

Instrument :

MSVOA\_W

ClientSampleId :

YBW38MS

Tgt Ion: 63 Resp: 270624

Ion Ratio Lower Upper

63 100





112 3.9 3.3 4.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/20/2023

Supervised By :Mahesh Dadoda 06/26/2023



#38

Bromodichloromethane

Concen: 22.859 ug/L

RT: 9.642 min Scan# 1272

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56




Tgt Ion: 83 Resp: 301232


Ion Ratio Lower Upper

83 100

85 64.9 44.2 82.0


127 7.7 6.2 9.2





#39  
cis-1,3-Dichloropropene  
Concen: 20.771 ug/L  
RT: 10.068 min Scan# 1342  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion: 75 Resp: 34621  
Ion Ratio Lower Upper  
75 100  
77 31.7 22.0 40.8

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#40  
4-Methyl-2-pentanone  
Concen: 60.137 ug/L  
RT: 10.209 min Scan# 1365  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Tgt Ion: 43 Resp: 402149  
Ion Ratio Lower Upper  
43 100  
58 36.6 29.0 43.4  
100 13.1 10.5 15.7





#42  
Toluene  
Concen: 21.573 ug/L  
RT: 10.385 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS



Tgt Ion: 91 Resp: 92404  
Ion Ratio Lower Upper  
91 100  
92 57.8 42.2 78.4




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#44  
trans-1,3-Dichloropropene  
Concen: 19.872 ug/L  
RT: 10.605 min Scan# 1430  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion: 75 Resp: 275951  
Ion Ratio Lower Upper  
75 100  
77 31.2 21.1 39.3





#45  
1,1,2-Trichloroethane  
Concen: 24.063 ug/L  
RT: 10.788 min Scan# 1459  
Delta R.T. 0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS






Tgt Ion: 97 Resp: 179879

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 99  | 64.5  | 43.3  | 80.5  |
| 83  | 88.8  | 62.9  | 116.9 |
| 85  | 59.4  | 40.1  | 74.5  |

Manual Integrations  
APPROVED




Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#46  
Tetrachloroethene  
Concen: 19.033 ug/L  
RT: 10.861 min Scan# 1472  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Tgt Ion:164 Resp: 144362


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 164 | 100   |       |       |
| 129 | 103.6 | 68.5  | 127.1 |
| 131 | 99.2  | 67.1  | 124.7 |
| 166 | 130.6 | 94.1  | 174.8 |





#48  
2-Hexanone  
Concen: 61.345 ug/L  
RT: 10.964 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS



Tgt Ion: 43 Resp: 289248  
Ion Ratio Lower Upper  
43 100  
58 50.4 42.2 63.2  
57 17.8 14.2 21.2  
100 10.2 8.3 12.5




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#49  
Dibromochloromethane  
Concen: 22.735 ug/L  
RT: 11.129 min Scan# 1516  
Delta R.T. 0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion:129 Resp: 178186  
Ion Ratio Lower Upper  
129 100  
127 74.5 56.9 105.7





#50  
1,2-Dibromoethane  
Concen: 21.891 ug/L  
RT: 11.233 min Scan# 1533  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion:107 Resp: 151520  
Ion Ratio Lower Upper  
107 100  
109 94.9 64.0 119.0  
188 2.6 2.6 3.8

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#51  
Chlorobenzene  
Concen: 19.160 ug/L  
RT: 11.653 min Scan# 1602  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Tgt Ion:112 Resp: 514253  
Ion Ratio Lower Upper  
112 100  
114 31.7 23.0 42.6  
77 63.5 50.9 76.3





#52  
Ethylbenzene  
Concen: 20.242 ug/L  
RT: 11.727 min Scan# 1614  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion: 91 Resp: 97803  
Ion Ratio Lower Upper  
91 100  
106 29.5 21.2 39.4

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#53  
m,p-Xylene  
Concen: 19.655 ug/L  
RT: 11.836 min Scan# 1632  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56



Tgt Ion:106 Resp: 353301  
Ion Ratio Lower Upper  
106 100  
91 201.2 149.5 277.7









#54  
o-Xylene  
Concen: 20.595 ug/L  
RT: 12.159 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS

Tgt Ion:106 Resp: 34940  
Ion Ratio Lower Upper  
106 100  
91 222.0 150.9 280.3




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#55  
Styrene  
Concen: 18.334 ug/L  
RT: 12.178 min Scan# 1688  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Tgt Ion:104 Resp: 533610  
Ion Ratio Lower Upper  
104 100  
78 49.0 32.6 60.6









#57  
1,1,2,2-Tetrachloroethane  
Concen: 25.073 ug/L  
RT: 12.714 min Scan# 1776  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS

Tgt Ion: 83 Resp: 22683  
Ion Ratio Lower Upper  
83 100  
85 63.5 42.6 79.0  
131 9.0 7.2 10.8




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#59  
Bromoform  
Concen: 28.641 ug/L  
RT: 12.348 min Scan# 1716  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion:173 Resp: 95791  
Ion Ratio Lower Upper  
173 100  
175 49.3 38.6 57.8  
254 8.8 7.4 11.0





#60  
Isopropylbenzene  
Concen: 26.963 ug/L  
RT: 12.464 min Scan# 1735  
Delta R.T. 0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion:105 Resp: 943192  
Ion Ratio Lower Upper  
105 100  
120 25.5 19.8 29.8  
77 16.8 13.3 19.9

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#61  
1,2,3-Trichloropropane  
Concen: 33.360 ug/L  
RT: 12.769 min Scan# 1785  
Delta R.T. 0.006 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56



Tgt Ion: 75 Resp: 166395  
Ion Ratio Lower Upper  
75 100  
77 31.9 25.2 37.8  
110 35.4 28.7 43.1









#62  
1,3,5-Trimethylbenzene  
Concen: 25.664 ug/L  
RT: 12.940 min Scan# 1813  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56

Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS

Tgt Ion:105 Resp: 71366  
Ion Ratio Lower Upper  
105 100  
120 47.3 38.8 58.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#63  
1,2,4-Trimethylbenzene  
Concen: 24.250 ug/L  
RT: 13.245 min Scan# 1863  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56


Tgt Ion:105 Resp: 617637  
Ion Ratio Lower Upper  
105 100  
120 44.3 36.2 54.2





#64  
1,3-Dichlorobenzene  
Concen: 19.176 ug/L  
RT: 13.495 min Scan# 1904  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56




Instrument :  
MSVOA\_W  
ClientSampleId :  
YBW38MS




Tgt Ion:146 Resp: 28204  
Ion Ratio Lower Upper  
146 100  
111 41.4 29.7 55.1  
148 59.4 45.4 84.4

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023




#65  
1,4-Dichlorobenzene  
Concen: 19.169 ug/L  
RT: 13.574 min Scan# 1917  
Delta R.T. -0.000 min  
Lab File: VW026254.D  
Acq: 19 Jun 2023 17:56



Tgt Ion:146 Resp: 283847  
Ion Ratio Lower Upper  
146 100  
111 41.0 29.2 54.2  
148 58.7 42.8 79.6





#67

1,2-Dichlorobenzene

Concen: 20.759 ug/L

RT: 13.866 min Scan# 1965

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56

Instrument :

MSVOA\_W

ClientSampleId :

YBW38MS





|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 146   | Resp: | 27417 |
| Ion Ratio | Lower | Upper |       |
| 146       | 100   |       |       |
| 111       | 43.0  | 35.4  | 53.0  |
| 148       | 64.6  | 51.7  | 77.5  |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/20/2023

Supervised By :Mahesh Dadoda 06/26/2023



#68

1,2-Dibromo-3-chloropropane

Concen: 28.994 ug/L




RT: 14.482 min Scan# 2066


Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 75    | Resp: | 33506 |
| Ion Ratio | Lower | Upper |       |
| 75        | 100   |       |       |
| 155       | 64.4  | 59.0  | 88.6  |
| 157       | 78.2  | 71.7  | 107.5 |





#69

1,3,5-Trichlorobenzene

Concen: 14.219 ug/L

RT: 14.622 min Scan# 2089

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56

Instrument :

MSVOA\_W

ClientSampleId :

YBW38MS

Tgt Ion:180 Resp: 14365

Ion Ratio Lower Upper

180 100

182 94.4 77.0 115.4

184 27.9 23.7 35.5





145 30.5 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/20/2023

Supervised By :Mahesh Dadoda 06/26/2023



#70

1,2,4-trichlorobenzene

Concen: 10.441 ug/L

RT: 15.128 min Scan# 2172

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56




Tgt Ion:180 Resp: 87044


Ion Ratio Lower Upper

180 100

182 96.2 78.9 118.3

145 33.8 27.2 40.8





#71

Naphthalene

Concen: 14.995 ug/L

RT: 15.360 min Scan# 2210

Delta R.T. -0.000 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56

Instrument :

MSVOA\_W

ClientSampleId :

YBW38MS





Tgt Ion:128 Resp: 22281

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 128 | 100   |       |       |
| 127 | 13.8  | 10.6  | 15.8  |
| 129 | 10.7  | 8.8   | 13.2  |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/20/2023  
Supervised By :Mahesh Dadoda 06/26/2023



#72

1,2,3-Trichlorobenzene

Concen: 11.936 ug/L

RT: 15.549 min Scan# 2241

Delta R.T. 0.006 min

Lab File: VW026254.D

Acq: 19 Jun 2023 17:56

Tgt Ion:180 Resp: 86964

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 180 | 100   |       |       |
| 182 | 92.8  | 76.3  | 114.5 |
| 145 | 34.1  | 27.0  | 40.6  |

