

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061924\
 Data File : VW029235.D
 Acq On : 20 Jun 2024 00:16
 Operator : SY/MD
 Sample : P2963-12
 Misc : 4.59g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CR6

Quant Time: Jun 20 01:57:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 20 00:39:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	408105	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	337009	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.549	152	111343	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	107947	19.110	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.440%
7) Chloroethane-d5	2.899	69	97236	20.384	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.520%
11) 1,1-Dichloroethene-d2	4.021	65	50897	18.662	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.640%
21) 2-Butanone-d5	7.075	46	70106	41.483	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.960%
24) Chloroform-d	7.648	84	258904	21.267	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.080%
26) 1,2-Dichloroethane-d4	8.307	65	147014	21.033	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.120%
32) Benzene-d6	8.276	84	479780	23.438	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.760%
36) 1,2-Dichloropropane-d6	9.270	67	162386	24.956	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.840%
41) Toluene-d8	10.318	98	393290	21.716	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.880%
43) trans-1,3-Dichloroprop...	10.574	79	45533	17.365	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.440%
47) 2-Hexanone-d5	10.922	63	40787	39.985	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.980%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	115690	22.339	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.360%
66) 1,2-Dichlorobenzene-d4	13.848	152	94156	24.362	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061924\
Data File : VW029235.D
Acq On : 20 Jun 2024 00:16
Operator : SY/MD
Sample : P2963-12
Misc : 4.59g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CR6

Quant Time: Jun 20 01:57:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 20 00:39:34 2024
Response via : Initial Calibration

