

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061924\
 Data File : VW029241.D
 Acq On : 20 Jun 2024 02:40
 Operator : SY/MD
 Sample : P2963-18
 Misc : 4.40g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CS9

Quant Time: Jun 20 05:06:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 20 00:39:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	394813	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	323930	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	110216	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	100181	18.332	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.320%
7) Chloroethane-d5	2.899	69	90380	19.585	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.320%
11) 1,1-Dichloroethene-d2	4.021	65	46847	17.755	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.040%
21) 2-Butanone-d5	7.075	46	59608	36.459	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.920%
24) Chloroform-d	7.648	84	232165	19.713	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.840%
26) 1,2-Dichloroethane-d4	8.307	65	131876	19.502	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.000%
32) Benzene-d6	8.270	84	428281	21.767	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.080%
36) 1,2-Dichloropropane-d6	9.270	67	144660	23.129	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.520%
41) Toluene-d8	10.319	98	351079	20.168	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.680%
43) trans-1,3-Dichloroprop...	10.575	79	39848	15.810	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	63.240%
47) 2-Hexanone-d5	10.922	63	34373	35.058	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.120%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	101371	20.364	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.440%
66) 1,2-Dichlorobenzene-d4	13.848	152	84181	22.004	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061924\
Data File : VW029241.D
Acq On : 20 Jun 2024 02:40
Operator : SY/MD
Sample : P2963-18
Misc : 4.40g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CS9

Quant Time: Jun 20 05:06:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 20 00:39:34 2024
Response via : Initial Calibration

