

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061924\
 Data File : VW029204.D
 Acq On : 19 Jun 2024 10:57
 Operator : SY/MD
 Sample : VIBLK461
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK461

Quant Time: Jun 20 01:54:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 20 00:39:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	247521	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	233108	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	100745	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	76364	22.290	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.160%
7) Chloroethane-d5	2.899	69	61688	21.322	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.280%
11) 1,1-Dichloroethene-d2	4.021	65	34715	20.986	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.960%
21) 2-Butanone-d5	7.075	46	43798	42.730	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.460%
24) Chloroform-d	7.648	84	158337	21.444	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.760%
26) 1,2-Dichloroethane-d4	8.306	65	94887	22.383	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.520%
32) Benzene-d6	8.270	84	312106	22.043	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.160%
36) 1,2-Dichloropropane-d6	9.270	67	100599	22.351	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.400%
41) Toluene-d8	10.324	98	279618	22.321	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.280%
43) trans-1,3-Dichloroprop...	10.574	79	38369	21.155	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.600%
47) 2-Hexanone-d5	10.922	63	29910	42.392	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.780%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	75082	20.960	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.840%
66) 1,2-Dichlorobenzene-d4	13.848	152	86413	24.711	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061924\
Data File : VW029204.D
Acq On : 19 Jun 2024 10:57
Operator : SY/MD
Sample : VIBLK461
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK461

Quant Time: Jun 20 01:54:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 20 00:39:34 2024
Response via : Initial Calibration

