

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062024\
 Data File : VW029277.D
 Acq On : 20 Jun 2024 23:43
 Operator : SY/MD
 Sample : VIBLK470
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK470

Quant Time: Jun 21 04:51:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 02:43:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	284987	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	266176	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	122729	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	79545	20.166	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.680%
7) Chloroethane-d5	2.899	69	64402	19.334	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.320%
11) 1,1-Dichloroethene-d2	4.027	65	39294	20.632	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.520%
21) 2-Butanone-d5	7.075	46	51939	44.011	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.020%
24) Chloroform-d	7.648	84	169288	19.913	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.640%
26) 1,2-Dichloroethane-d4	8.307	65	101105	20.714	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.840%
32) Benzene-d6	8.270	84	343477	21.245	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.960%
36) 1,2-Dichloropropane-d6	9.270	67	108699	21.150	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.600%
41) Toluene-d8	10.325	98	303971	21.251	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	10.575	79	40425	19.519	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.080%
47) 2-Hexanone-d5	10.922	63	34239	42.499	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	83669	20.455	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.840%
66) 1,2-Dichlorobenzene-d4	13.848	152	98572	23.139	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.560%
Target Compounds						
16) Methylene chloride	4.917	84	33276	6.640	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062024\
Data File : VW029277.D
Acq On : 20 Jun 2024 23:43
Operator : SY/MD
Sample : VIBLK470
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK470

Quant Time: Jun 21 04:51:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 02:43:08 2024
Response via : Initial Calibration

