

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062024\
 Data File : VW029263.D
 Acq On : 20 Jun 2024 17:18
 Operator : SY/MD
 Sample : P2964-03
 Misc : 5.19g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4B50

Quant Time: Jun 21 01:33:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 01:26:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	392110	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	296758	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	86499	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	89117	16.420	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.680%
7) Chloroethane-d5	2.893	69	78629	17.156	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.640%
11) 1,1-Dichloroethene-d2	4.021	65	44770	17.085	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.320%
21) 2-Butanone-d5	7.075	46	81109	49.952	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.900%
24) Chloroform-d	7.642	84	227755	19.472	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.880%
26) 1,2-Dichloroethane-d4	8.301	65	140065	20.856	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.440%
32) Benzene-d6	8.270	84	425568	23.610	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.440%
36) 1,2-Dichloropropane-d6	9.270	67	143562	25.055	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.240%
41) Toluene-d8	10.319	98	319645	20.044	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.160%
43) trans-1,3-Dichloroprop...	10.575	79	43190	18.705	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.840%
47) 2-Hexanone-d5	10.922	63	54951	61.178	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.360%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	106889	23.439	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	58484	19.479	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	77.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062024\
Data File : VW029263.D
Acq On : 20 Jun 2024 17:18
Operator : SY/MD
Sample : P2964-03
Misc : 5.19g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4B50

Quant Time: Jun 21 01:33:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 01:26:21 2024
Response via : Initial Calibration

