

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062118\
 Data File : VW003382.D
 Acq On : 21 Jun 2018 12:53
 Operator : SY/AP
 Sample : J3569-07
 Misc : 6.12G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DAXQ8

Quant Time: Jun 22 05:07:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 22 04:53:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	348331	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.64	117	287246	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	103677	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	119842	25.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.48%
7) Chloroethane-d5	2.89	69	96429	24.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.84%
10) 1,1-Dichloroethene-d2	4.00	63	175986	18.76	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.04%
20) 2-Butanone-d5	7.10	46	72478	48.69	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.38%
24) Chloroform-d	7.65	84	224732	24.53	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.12%
26) 1,2-Dichloroethane-d4	8.31	65	141603	26.01	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.04%
29) Benzene-d6	8.27	84	429466	28.18	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	112.72%
33) 1,2-Dichloropropane-d6	9.28	67	138245	28.65	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	114.60%
37) Toluene-d8	10.33	98	381007	26.15	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	104.60%
38) trans-1,3-Dichloropropene-	10.58	79	59492	26.27	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	105.08%
39) 2-Hexanone-d5	10.93	63	52862	53.61	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.22%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	103623	23.09	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.36%
61) 1,2-Dichlorobenzene-d4	13.87	152	91387	23.34	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.36%

Target Compounds					Ovalue
13) Acetone	4.15	43	150098	94.60 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062118\
Data File : VW003382.D
Acq On : 21 Jun 2018 12:53
Operator : SY/AP
Sample : J3569-07
Misc : 6.12G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DAXQ8

Quant Time: Jun 22 05:07:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 22 04:53:24 2018
Response via : Initial Calibration

