

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062118\
 Data File : VW003392.D
 Acq On : 21 Jun 2018 17:11
 Operator : SY/AP
 Sample : J3619-05
 Misc : 6.50G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DAZ69

Quant Time: Jun 22 04:56:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 22 04:53:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	299294	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	288125	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	134571	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	94752	23.34	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.36%
7) Chloroethane-d5	2.89	69	73919	21.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.32%
10) 1,1-Dichloroethene-d2	4.01	63	139584	17.32	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.28%
20) 2-Butanone-d5	7.09	46	81616	63.81	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	127.62%
24) Chloroform-d	7.65	84	188307	23.92	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.68%
26) 1,2-Dichloroethane-d4	8.31	65	125309	26.78	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.12%
29) Benzene-d6	8.27	84	359947	23.54	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.16%
33) 1,2-Dichloropropane-d6	9.28	67	117960	24.38	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.52%
37) Toluene-d8	10.33	98	336334	23.01	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.04%
38) trans-1,3-Dichloropropene-	10.59	79	53499	23.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.20%
39) 2-Hexanone-d5	10.93	63	60403	61.07	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.14%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	121042	26.88	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.52%
61) 1,2-Dichlorobenzene-d4	13.86	152	129429	25.46	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.84%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062118\
Data File : VW003392.D
Acq On : 21 Jun 2018 17:11
Operator : SY/AP
Sample : J3619-05
Misc : 6.50G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DAZ69

Quant Time: Jun 22 04:56:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 22 04:53:24 2018
Response via : Initial Calibration

