

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062218\
 Data File : VW003418.D
 Acq On : 22 Jun 2018 10:57
 Operator : SY/AP
 Sample : VW0622SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK90

Quant Time: Jun 23 02:15:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 23 02:14:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	351324	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.64	117	330313	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	153738	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	93864	19.70	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.80%
7) Chloroethane-d5	2.89	69	77026	19.38	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.52%
10) 1,1-Dichloroethene-d2	4.01	63	138624	14.65	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.60%
20) 2-Butanone-d5	7.10	46	71712	47.76	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.52%
24) Chloroform-d	7.65	84	187927	20.33	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.32%
26) 1,2-Dichloroethane-d4	8.31	65	115397	21.01	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.04%
29) Benzene-d6	8.27	84	354563	20.23	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.92%
33) 1,2-Dichloropropane-d6	9.28	67	113015	20.37	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.48%
37) Toluene-d8	10.33	98	320189	19.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.44%
38) trans-1,3-Dichloropropene-	10.59	79	49010	18.82	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.28%
39) 2-Hexanone-d5	10.93	63	50866	44.86	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.72%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	107918	20.91	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.64%
61) 1,2-Dichlorobenzene-d4	13.87	152	124404	21.42	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.68%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062218\
Data File : VW003418.D
Acq On : 22 Jun 2018 10:57
Operator : SY/AP
Sample : VW0622SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK90

Quant Time: Jun 23 02:15:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 23 02:14:03 2018
Response via : Initial Calibration

