

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062218\
 Data File : VW003431.D
 Acq On : 22 Jun 2018 17:29
 Operator : SY/AP
 Sample : J3621-09
 Misc : 6.97G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DAZ61

Quant Time: Jun 23 02:16:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 23 02:14:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	322210	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	308080	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	136160	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	90469	20.70	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.80%
7) Chloroethane-d5	2.89	69	75411	20.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.72%
10) 1,1-Dichloroethene-d2	4.01	63	141474	16.30	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.20%
20) 2-Butanone-d5	7.09	46	55437	40.26	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.52%
24) Chloroform-d	7.65	84	195598	23.08	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.32%
26) 1,2-Dichloroethane-d4	8.31	65	115195	22.87	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.48%
29) Benzene-d6	8.28	84	358822	21.95	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.80%
33) 1,2-Dichloropropane-d6	9.28	67	120642	23.32	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.28%
37) Toluene-d8	10.33	98	330928	21.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.68%
38) trans-1,3-Dichloropropene-	10.58	79	48585	20.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.00%
39) 2-Hexanone-d5	10.93	63	40699	38.48	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.96%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	97940	20.34	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.36%
61) 1,2-Dichlorobenzene-d4	13.87	152	123224	23.96	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.84%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062218\
Data File : VW003431.D
Acq On : 22 Jun 2018 17:29
Operator : SY/AP
Sample : J3621-09
Misc : 6.97G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DAZ61

Quant Time: Jun 23 02:16:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 23 02:14:03 2018
Response via : Initial Calibration

