

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062418\
 Data File : VW003507.D
 Acq On : 24 Jun 2018 13:47
 Operator : SY/AP
 Sample : J3669-09
 Misc : 6.23G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WP021SB484-39-41-180620

Quant Time: Jun 24 23:46:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	152740	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	250274	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	231955	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	109450	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	107412	64.92	ug/l	0.00
Spiked Amount			Recovery	=	129.84%	
35) Dibromofluoromethane	7.88	113	89026	57.25	ug/l	0.00
Spiked Amount			Recovery	=	114.50%	
50) Toluene-d8	10.33	98	303018	49.14	ug/l	0.00
Spiked Amount			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	12.63	95	107318	49.18	ug/l	0.00
Spiked Amount			Recovery	=	98.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	4.05	101	3624	2.56	ug/l #	73
12) 1,1-Dichloroethene	4.04	96	32862	23.77	ug/l	90
16) Acetone	4.14	43	6338	18.27	ug/l #	88
20) Methylene Chloride	4.92	84	7103	1.24	ug/l	93
27) cis-1,2-Dichloroethene	7.17	96	454346	274.85	ug/l	92
30) Chloroform	7.68	83	3896	1.34	ug/l	97
44) Trichloroethene	9.10	130	2688381	1522.85	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062418\
Data File : VW003507.D
Acq On : 24 Jun 2018 13:47
Operator : SY/AP
Sample : J3669-09
Misc : 6.23G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WP021SB484-39-41-180620

Quant Time: Jun 24 23:46:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

