

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062424\
 Data File : VW029323.D
 Acq On : 24 Jun 2024 13:34
 Operator : SY/MD
 Sample : VIBLK474
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK474

Quant Time: Jun 25 01:39:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 25 01:37:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	307636	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	283757	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	124692	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	91599	21.512	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.040%
7) Chloroethane-d5	2.899	69	70813	19.693	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.760%
11) 1,1-Dichloroethene-d2	4.021	65	42704	20.771	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.080%
21) 2-Butanone-d5	7.075	46	47330	37.153	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.300%
24) Chloroform-d	7.648	84	182188	19.853	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.400%
26) 1,2-Dichloroethane-d4	8.307	65	104213	19.779	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	79.120%
32) Benzene-d6	8.276	84	370405	21.491	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.960%
36) 1,2-Dichloropropane-d6	9.276	67	114094	20.825	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.280%
41) Toluene-d8	10.325	98	329437	21.604	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.400%
43) trans-1,3-Dichloroprop...	10.575	79	43488	19.697	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.800%
47) 2-Hexanone-d5	10.922	63	30302	35.281	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.560%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	81751	18.748	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	75.000%
66) 1,2-Dichlorobenzene-d4	13.848	152	99995	23.103	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.400%
Target Compounds						
16) Methylene chloride	4.923	84	14188	2.623	ug/L	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062424\
Data File : VW029323.D
Acq On : 24 Jun 2024 13:34
Operator : SY/MD
Sample : VIBLK474
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK474

Quant Time: Jun 25 01:39:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 25 01:37:56 2024
Response via : Initial Calibration

