

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062424\
 Data File : VW029321.D
 Acq On : 24 Jun 2024 12:46
 Operator : SY/MD
 Sample : P2982-18
 Misc : 6.31g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BB9

Quant Time: Jun 25 01:39:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 25 01:37:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	460482	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	410441	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	166866	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	123807	19.425	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.680%
7) Chloroethane-d5	2.899	69	110587	20.546	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.200%
11) 1,1-Dichloroethene-d2	4.021	65	60367	19.616	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.480%
21) 2-Butanone-d5	7.075	46	72805	38.180	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.360%
24) Chloroform-d	7.648	84	303858	22.121	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.480%
26) 1,2-Dichloroethane-d4	8.307	65	165779	21.020	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.080%
32) Benzene-d6	8.270	84	578971	23.224	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.880%
36) 1,2-Dichloropropane-d6	9.270	67	188471	23.782	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.120%
41) Toluene-d8	10.319	98	489331	22.185	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.760%
43) trans-1,3-Dichloroprop...	10.575	79	64742	20.273	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.080%
47) 2-Hexanone-d5	10.922	63	51503	41.458	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.920%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	127422	20.202	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.800%
66) 1,2-Dichlorobenzene-d4	13.849	152	139510	24.086	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062424\
Data File : VW029321.D
Acq On : 24 Jun 2024 12:46
Operator : SY/MD
Sample : P2982-18
Misc : 6.31g/10mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BB9

Quant Time: Jun 25 01:39:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 25 01:37:56 2024
Response via : Initial Calibration

