

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062524\
 Data File : VW029346.D
 Acq On : 25 Jun 2024 13:40
 Operator : SY/MD
 Sample : VIBLK478
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK478

Quant Time: Jun 26 02:27:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 26 02:25:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	279318	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	271783	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	113313	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	95239	24.634	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	98.520%		
7) Chloroethane-d5	2.899	69	73103	22.391	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	89.560%		
11) 1,1-Dichloroethene-d2	4.021	65	44286	23.725	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	94.880%		
21) 2-Butanone-d5	7.075	46	52976	45.801	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	91.600%		
24) Chloroform-d	7.648	84	186509	22.384	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	89.520%		
26) 1,2-Dichloroethane-d4	8.307	65	109492	22.888	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	91.560%		
32) Benzene-d6	8.270	84	371111	22.481	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	89.920%		
36) 1,2-Dichloropropane-d6	9.270	67	118170	22.519	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	90.080%		
41) Toluene-d8	10.325	98	332488	22.765	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	91.040%		
43) trans-1,3-Dichloroprop...	10.575	79	44336	20.966	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	83.880%		
47) 2-Hexanone-d5	10.922	63	35038	42.593	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	85.180%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	87755	21.012	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	84.040%		
66) 1,2-Dichlorobenzene-d4	13.848	152	102996	26.186	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	104.760%		
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	13851	2.820	ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062524\
Data File : VW029346.D
Acq On : 25 Jun 2024 13:40
Operator : SY/MD
Sample : VIBLK478
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK478

Quant Time: Jun 26 02:27:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 26 02:25:08 2024
Response via : Initial Calibration





#16
Methylene chloride
Concen: 2.820 ug/L
RT: 4.917 min Scan# 49
Delta R.T. -0.000 min
Lab File: VW029346.D
Acq: 25 Jun 2024 13:40

Instrument :
MSVOA_W
ClientSampleId :
VIBLK478

Tgt Ion: 84 Resp: 13851
Ion Ratio Lower Upper
84 100
86 66.9 41.4 77.0
49 143.5 95.7 177.7

