

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062623\
 Data File : VW026337.D
 Acq On : 26 Jun 2023 15:47
 Operator : SY/MD
 Sample : 03383-04
 Misc : 5.04g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jun 27 01:53:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 27 01:51:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	633642	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	584286	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	246327	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	250746	27.239	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.960%
7) Chloroethane-d5	2.899	69	160503	24.527	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.120%
11) 1,1-Dichloroethene-d2	4.027	65	109498	26.274	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	105.080%
21) 2-Butanone-d5	7.081	46	146755	50.726	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.460%
24) Chloroform-d	7.648	84	424033	23.616	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.480%
26) 1,2-Dichloroethane-d4	8.410	65	369	0.037	ug/L	0.10
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.160%#
32) Benzene-d6	8.276	84	887755	24.547	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.200%
36) 1,2-Dichloropropane-d6	9.276	67	283327	24.109	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.440%
41) Toluene-d8	10.325	98	750351	23.703	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.800%
43) trans-1,3-Dichloroprop...	10.581	79	109291	23.163	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.640%
47) 2-Hexanone-d5	10.922	63	88258	48.901	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.800%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	225681	24.392	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.560%
66) 1,2-Dichlorobenzene-d4	13.848	152	231768	27.023	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	108.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062623\
Data File : VW026337.D
Acq On : 26 Jun 2023 15:47
Operator : SY/MD
Sample : 03383-04
Misc : 5.04g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jun 27 01:53:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062623SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 27 01:51:47 2023
Response via : Initial Calibration

