

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062624\
 Data File : VW029351.D
 Acq On : 26 Jun 2024 10:33
 Operator : SY/MD
 Sample : P3026-07
 Misc : 5.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Jun 27 05:36:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 27 05:32:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	262277	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	250543	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	111184	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	81957	22.576	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.320%
7) Chloroethane-d5	2.899	69	68320	22.286	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.160%
11) 1,1-Dichloroethene-d2	4.021	65	39699	22.649	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	90.600%
21) 2-Butanone-d5	7.081	46	50673	46.656	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.320%
24) Chloroform-d	7.648	84	180345	23.051	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	8.307	65	106517	23.712	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.840%
32) Benzene-d6	8.276	84	361802	23.775	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.080%
36) 1,2-Dichloropropane-d6	9.270	67	115579	23.892	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.560%
41) Toluene-d8	10.325	98	317270	23.565	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.240%
43) trans-1,3-Dichloroprop...	10.581	79	43143	22.132	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.520%
47) 2-Hexanone-d5	10.922	63	33283	43.890	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.780%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	89591	23.270	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.080%
66) 1,2-Dichlorobenzene-d4	13.849	152	104948	27.194	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	108.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062624\
Data File : VW029351.D
Acq On : 26 Jun 2024 10:33
Operator : SY/MD
Sample : P3026-07
Misc : 5.00g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Jun 27 05:36:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 27 05:32:27 2024
Response via : Initial Calibration

