

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062723\
 Data File : VW026345.D
 Acq On : 27 Jun 2023 11:26
 Operator : SY/MD
 Sample : 03421-04
 Misc : 5.86g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4642

Quant Time: Jun 28 01:20:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 28 01:18:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	916526	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	622390	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	157919	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	270373	20.306	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.240%
7) Chloroethane-d5	2.893	69	206406	21.806	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.240%
11) 1,1-Dichloroethene-d2	4.021	65	128112	21.252	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.000%
21) 2-Butanone-d5	7.075	46	162936	38.936	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.880%
24) Chloroform-d	7.648	84	511138	19.681	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.720%
26) 1,2-Dichloroethane-d4	8.307	65	309049	21.178	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.720%
32) Benzene-d6	8.276	84	1027782	26.679	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	106.720%
36) 1,2-Dichloropropane-d6	9.276	67	328736	26.261	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.040%
41) Toluene-d8	10.325	98	809582	24.009	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.040%
43) trans-1,3-Dichloroprop...	10.575	79	110756	22.036	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.160%
47) 2-Hexanone-d5	10.922	63	97459	50.693	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.380%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	186000	18.872	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	75.480%
66) 1,2-Dichlorobenzene-d4	13.848	152	102803	18.697	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	74.800%#
Target Compounds						
					Qvalue	
13) Acetone	4.118	43	260923	77.901	ug/L	99
22) 2-Butanone	7.173	43	100359	19.583	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062723\
Data File : VW026345.D
Acq On : 27 Jun 2023 11:26
Operator : SY/MD
Sample : 03421-04
Misc : 5.86g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4642

Quant Time: Jun 28 01:20:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062623SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 28 01:18:48 2023
Response via : Initial Calibration









#13
Acetone
Concen: 77.901 ug/L
RT: 4.118 min Scan# 366
Delta R.T. -0.012 min
Lab File: VW026345.D
Acq: 27 Jun 2023 11:26

Instrument :
MSVOA_W
ClientSampleId :
A4642

Tgt Ion: 43 Resp: 260923
Ion Ratio Lower Upper
43 100
58 28.5 0.0 55.8



#22
2-Butanone
Concen: 19.583 ug/L
RT: 7.173 min Scan# 867
Delta R.T. -0.000 min
Lab File: VW026345.D
Acq: 27 Jun 2023 11:26

Tgt Ion: 43 Resp: 100359
Ion Ratio Lower Upper
43 100
72 22.4 11.8 35.3

