

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063023\
 Data File : VW026476.D
 Acq On : 30 Jun 2023 16:41
 Operator : SY/MD
 Sample : 03478-12
 Misc : 5.23g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 YBW59

Quant Time: Jun 30 23:42:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 30 23:38:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	1112954	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	920112	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	379870	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	250891	19.799	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.200%
7) Chloroethane-d5	2.893	69	203651	19.736	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.960%
11) 1,1-Dichloroethene-d2	4.021	65	130440	20.573	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.280%
21) 2-Butanone-d5	7.075	46	225177	47.019	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.040%
24) Chloroform-d	7.648	84	584302	21.539	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.160%
26) 1,2-Dichloroethane-d4	8.307	65	340174	22.942	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.760%
32) Benzene-d6	8.276	84	1266000	23.658	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.640%
36) 1,2-Dichloropropane-d6	9.270	67	414895	24.326	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.320%
41) Toluene-d8	10.325	98	1063709	22.503	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.000%
43) trans-1,3-Dichloroprop...	10.575	79	151691	22.065	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.280%
47) 2-Hexanone-d5	10.922	63	153243	53.727	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.460%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	308519	23.428	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.720%
66) 1,2-Dichlorobenzene-d4	13.848	152	286361	21.649	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.600%
Target Compounds					Qvalue	
13) Acetone	4.125	43	13546	4.004	ug/L	95
34) Trichloroethene	9.093	95	25818	1.759	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063023\
Data File : VW026476.D
Acq On : 30 Jun 2023 16:41
Operator : SY/MD
Sample : 03478-12
Misc : 5.23g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
YBW59

Quant Time: Jun 30 23:42:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062923SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 30 23:38:58 2023
Response via : Initial Calibration

