

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063023\
 Data File : VW026480.D
 Acq On : 30 Jun 2023 18:18
 Operator : SY/MD
 Sample : 03478-01
 Misc : 6.16g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 YBW57

Quant Time: Jun 30 23:43:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 30 23:38:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	962964	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	798073	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	329034	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	191218	17.440	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	69.760%	
7) Chloroethane-d5	2.893	69	153868	17.234	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	68.920%	
11) 1,1-Dichloroethene-d2	4.027	65	99285	18.098	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	72.400%	
21) 2-Butanone-d5	7.075	46	209778	50.626	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	101.260%	
24) Chloroform-d	7.648	84	388725	16.561	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	66.240%	
26) 1,2-Dichloroethane-d4	8.307	65	278869	21.737	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	86.960%	
32) Benzene-d6	8.276	84	971909	20.940	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	83.760%	
36) 1,2-Dichloropropane-d6	9.270	67	324878	21.961	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	87.840%	
41) Toluene-d8	10.325	98	816319	19.910	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	79.640%	
43) trans-1,3-Dichloroprop...	10.575	79	123634	20.734	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	82.920%	
47) 2-Hexanone-d5	10.922	63	130720	52.839	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	105.680%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	196882	17.237	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	68.960%	
66) 1,2-Dichlorobenzene-d4	13.848	152	234364	20.456	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	81.840%	
Target Compounds						
					Qvalue	
13) Acetone	4.118	43	20806	7.108	ug/L	100
25) Chloroform	7.673	83	71304	3.147	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063023\
Data File : VW026480.D
Acq On : 30 Jun 2023 18:18
Operator : SY/MD
Sample : 03478-01
Misc : 6.16g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
YBW57

Quant Time: Jun 30 23:43:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062923SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 30 23:38:58 2023
Response via : Initial Calibration

