

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063023\
 Data File : VW026467.D
 Acq On : 30 Jun 2023 13:04
 Operator : SY/MD
 Sample : 03458-20
 Misc : 4.98g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jun 30 23:40:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 30 23:38:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	835749	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	727464	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	344654	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	227607	23.919	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.680%
7) Chloroethane-d5	2.893	69	176420	22.768	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.080%
11) 1,1-Dichloroethene-d2	4.015	65	115112	24.177	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	96.720%
21) 2-Butanone-d5	7.075	46	174635	48.560	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.120%
24) Chloroform-d	7.648	84	471197	23.131	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.520%
26) 1,2-Dichloroethane-d4	8.307	65	269855	24.236	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.960%
32) Benzene-d6	8.276	84	1033904	24.437	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.760%
36) 1,2-Dichloropropane-d6	9.270	67	328135	24.334	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.320%
41) Toluene-d8	10.319	98	885473	23.693	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.760%
43) trans-1,3-Dichloroprop...	10.575	79	124141	22.840	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.360%
47) 2-Hexanone-d5	10.922	63	114366	50.715	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.440%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	261135	25.081	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.320%
66) 1,2-Dichlorobenzene-d4	13.848	152	293349	24.444	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063023\
Data File : VW026467.D
Acq On : 30 Jun 2023 13:04
Operator : SY/MD
Sample : 03458-20
Misc : 4.98g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jun 30 23:40:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM062923SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 30 23:38:58 2023
Response via : Initial Calibration

