


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW070219\
Data File : VW011099.D
Acq On : 02 Jul 2019 08:45
Operator : SY/VA
Sample : BFB22
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB22

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070219S.M
Title : VOC Analysis
Last Update : Wed Jul 03 05:51:41 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1750

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	16069	PASS
75	95	30	80	54.3	38146	PASS
95	95	100	100	100.0	70304	PASS
96	95	5	9	6.6	4625	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	75.3	52930	PASS
175	174	5	9	7.9	4189	PASS
176	174	95	101	95.1	50354	PASS
177	176	5	9	7.0	3528	PASS

m/z	Abundance
37.90	57.0
38.90	68.0
39.90	194.0
41.10	110.0
43.10	108.0
44.00	275.0
55.10	65.0
57.20	127.0
73.00	79.0
132.70	56.0

Instrument :
MSVOA_W
ClientSampleId :
BFB22

m/z	Abundance
36.00	949.0
37.10	5238.0
38.00	4396.0
39.00	1943.0
39.90	309.0
41.10	116.0
42.30	52.0
43.00	154.0
44.00	1008.0
45.00	1106.0
46.00	97.0
47.10	1178.0
48.10	614.0
49.00	3957.0
50.00	17952.0
51.00	5884.0
52.00	311.0
52.90	56.0
54.90	367.0
56.00	1619.0
57.00	2509.0
57.90	117.0
59.00	355.0
60.00	825.0
61.00	4390.0
62.00	3939.0
63.00	3325.0
64.00	394.0
64.90	334.0
66.00	56.0
67.00	284.0
68.00	8657.0
69.00	8857.0
70.00	726.0
71.20	97.0
72.00	445.0
73.00	6062.0
74.00	13940.0
75.00	41848.0
76.00	3741.0
77.00	540.0
78.10	309.0
78.90	2384.0
80.00	817.0
81.00	2281.0
81.90	553.0
83.20	65.0
84.80	84.0
86.20	221.0
87.00	3141.0
88.00	3113.0
89.20	58.0
90.90	287.0
92.00	2241.0
93.00	3307.0
94.00	9132.0
95.00	74936.0
96.00	5346.0
97.00	167.0
102.90	477.0
103.90	452.0
105.00	258.0
105.90	384.0
106.80	106.0
109.90	191.0
110.80	186.0
111.80	164.0
112.70	65.0
114.90	195.0
115.70	223.0
116.90	611.0
117.90	502.0
119.00	651.0
119.90	80.0
120.80	67.0
125.00	621.0
126.00	141.0
127.90	287.0
128.90	227.0
130.00	280.0
131.10	197.0
132.10	118.0
133.00	1788.0
134.90	239.0
136.90	134.0
140.90	706.0
142.00	70.0
142.80	986.0
144.00	68.0
144.90	139.0
145.90	145.0
147.00	221.0

Instrument :
MSVOA_W
ClientSampleId :
BFB22

147.70	204.0
148.80	111.0
149.00	118.0
149.80	55.0
152.90	105.0
153.80	55.0
154.90	180.0
156.80	144.0
158.70	85.0
160.90	154.0
162.80	376.0
163.90	112.0
164.90	234.0
170.70	83.0
171.90	413.0
173.90	52280.0
175.00	4210.0
175.90	49592.0
177.00	3660.0
178.00	333.0
178.90	393.0
180.10	75.0
189.00	216.0
191.00	1950.0
192.10	304.0
193.00	2523.0
193.90	371.0
194.90	262.0
202.90	277.0
204.90	489.0
206.10	192.0
207.10	1374.0
208.00	298.0
208.90	134.0
219.00	109.0
219.90	53.0
221.10	84.0
235.00	247.0
236.00	51.0
249.00	1305.0
250.00	398.0
251.00	583.0
252.00	153.0
252.90	116.0
265.00	1616.0
266.10	446.0
266.80	274.0
267.10	303.0
268.10	176.0
269.00	58.0
279.00	185.0
281.00	22632.0
282.00	6363.0
283.00	3576.0
283.90	610.0
285.00	311.0

Instrument :
MSVOA_W
ClientSampleId :
BFB22

m/z	Abundance
36.00	952.0
37.10	5075.0
38.00	4471.0
39.10	1801.0
39.90	295.0
41.10	125.0
42.30	56.0
43.00	145.0
44.10	904.0
45.00	1058.0
46.20	208.0
47.00	1181.0
48.00	523.0
49.00	3801.0
50.00	17736.0
51.00	5343.0
52.00	196.0
52.90	97.0
54.90	247.0
56.00	1291.0
57.00	2328.0
58.10	90.0
59.10	213.0
60.00	907.0
61.00	4432.0
62.00	4303.0
63.00	2995.0
64.00	381.0
65.00	171.0
67.20	298.0
68.00	8578.0
69.00	8866.0
70.10	852.0
72.10	528.0
73.00	4888.0
74.00	14137.0
75.00	42304.0
76.10	3643.0
77.10	493.0
78.10	325.0
79.00	2241.0
79.90	678.0
80.90	2359.0
82.00	607.0
86.10	130.0
86.90	3141.0
87.90	3095.0
89.20	64.0
90.90	326.0
92.00	1995.0
93.00	3231.0
94.00	9296.0
95.00	78320.0
96.00	5114.0
96.90	155.0
103.00	260.0
104.00	296.0
104.80	124.0
105.90	364.0
106.90	100.0
110.00	136.0
110.70	184.0
111.80	92.0
112.80	76.0
114.80	147.0
116.00	279.0
116.90	640.0
118.00	491.0
118.80	574.0
119.90	70.0
121.00	73.0
124.90	543.0
126.00	97.0
127.70	204.0
128.80	194.0
129.90	249.0
130.90	92.0
133.00	1053.0
133.90	326.0
134.90	242.0
136.80	112.0
140.90	810.0
142.00	196.0
142.90	827.0
144.90	97.0
145.90	147.0
147.00	153.0
147.80	176.0
149.00	107.0
149.70	75.0
152.90	114.0
154.90	214.0

Instrument :
MSVOA_W
ClientSampleId :
BFB22

157.00	156.0
158.70	107.0
159.20	58.0
160.80	152.0
163.00	307.0
163.80	134.0
164.70	148.0
165.20	83.0
170.80	103.0
171.90	396.0
174.00	58688.0
174.90	4770.0
175.90	55464.0
176.90	3691.0
178.10	256.0
179.00	443.0
180.10	87.0
189.00	128.0
190.10	73.0
191.00	861.0
192.00	283.0
193.00	1328.0
193.90	289.0
194.90	190.0
202.80	148.0
204.90	312.0
205.70	71.0
207.00	949.0
208.10	175.0
209.10	186.0
234.80	137.0
236.00	64.0
249.00	971.0
249.90	288.0
251.00	435.0
251.90	93.0
265.00	1163.0
265.90	269.0
267.00	327.0
268.10	66.0
278.90	123.0
281.10	16912.0
282.10	4548.0
283.00	2886.0
284.00	557.0
285.00	153.0

Instrument :
MSVOA_W
ClientSampleId :
BFB22

m/z	Abundance
36.10	683.0
37.10	4085.0
38.00	3115.0
39.10	1262.0
40.10	352.0
41.10	99.0
43.10	97.0
44.00	689.0
45.00	623.0
46.00	146.0
47.10	857.0
48.10	365.0
49.00	2871.0
50.00	12519.0
51.00	4071.0
52.00	214.0
52.90	71.0
55.10	139.0
56.00	911.0
57.00	1502.0
57.90	92.0
59.00	110.0
60.00	716.0
61.00	3269.0
62.00	2818.0
63.00	2204.0
63.90	231.0
65.00	220.0
67.00	282.0
68.00	6273.0
69.00	6732.0
70.00	363.0
72.00	486.0
73.00	3164.0
74.00	10627.0
75.00	30288.0
76.10	2545.0
77.00	439.0
77.80	267.0
78.90	1652.0
79.90	586.0
80.90	1640.0
81.90	427.0
85.70	128.0
86.90	2095.0
87.90	2169.0
88.70	115.0
91.00	318.0
92.00	1864.0
92.90	2095.0
94.00	6813.0
95.00	57656.0
96.00	3415.0
97.10	141.0
102.70	105.0
103.10	84.0
103.90	319.0
105.00	111.0
106.00	207.0
106.80	100.0
109.90	78.0
110.80	95.0
112.70	86.0
114.90	136.0
116.00	214.0
116.80	456.0
117.90	371.0
118.90	277.0
125.00	113.0
126.00	81.0
128.00	193.0
129.00	110.0
129.80	260.0
131.00	128.0
133.00	534.0
134.00	104.0
134.90	91.0
136.80	117.0
141.00	601.0
142.90	726.0
145.10	94.0
146.00	75.0
146.70	181.0
147.80	162.0
149.70	82.0
152.80	56.0
154.90	207.0
157.00	147.0
158.80	89.0
159.20	60.0
160.90	100.0
162.90	125.0

Instrument :
MSVOA_W
ClientSampleId :
BFB22

171.90	332.0
173.90	47824.0
175.00	3587.0
175.90	46008.0
177.00	3235.0
178.00	125.0
179.10	133.0
180.10	53.0
189.00	84.0
191.00	458.0
192.10	132.0
193.00	726.0
194.00	148.0
194.90	110.0
202.90	79.0
204.90	150.0
207.10	544.0
208.10	108.0
234.90	64.0
235.00	65.0
249.00	497.0
249.90	156.0
250.80	221.0
251.10	231.0
264.90	495.0
266.20	180.0
267.10	152.0
267.90	73.0
279.10	132.0
281.10	8134.0
282.00	2222.0
283.10	1219.0
284.10	290.0
284.80	56.0
285.20	57.0

Instrument :
MSVOA_W
ClientSampleId :
BFB22