

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070224\
 Data File : VW029447.D
 Acq On : 02 Jul 2024 16:19
 Operator : SY/MD
 Sample : P3100-12
 Misc : 4.74g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BX0

Quant Time: Jul 03 05:44:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 03 05:41:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	272408	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	170211	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	33125	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	87330	23.162	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.640%
7) Chloroethane-d5	2.893	69	78260	24.579	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.320%
11) 1,1-Dichloroethene-d2	4.015	65	42074	23.112	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	92.440%
21) 2-Butanone-d5	7.075	46	36002	31.915	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	63.840%
24) Chloroform-d	7.648	84	193038	23.756	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.040%
26) 1,2-Dichloroethane-d4	8.307	65	112307	24.071	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.280%
32) Benzene-d6	8.276	84	356120	34.446	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	137.800%#
36) 1,2-Dichloropropane-d6	9.270	67	118311	36.000	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	144.000%#
41) Toluene-d8	10.319	98	246759	26.977	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	107.920%
43) trans-1,3-Dichloroprop...	10.575	79	28604	21.598	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.400%
47) 2-Hexanone-d5	10.922	63	23597	45.803	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.600%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	57514	21.988	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.960%
66) 1,2-Dichlorobenzene-d4	13.849	152	27700	24.091	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.360%
Target Compounds					Qvalue	
13) Acetone	4.131	43	4188	4.369	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070224\
Data File : VW029447.D
Acq On : 02 Jul 2024 16:19
Operator : SY/MD
Sample : P3100-12
Misc : 4.74g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BX0

Quant Time: Jul 03 05:44:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 03 05:41:51 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 4.369 ug/L
 RT: 4.131 min Scan# 30
 Delta R.T. 0.012 min
 Lab File: VW029447.D
 Acq: 02 Jul 2024 16:19

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BX0

Tgt Ion: 43 Resp: 4188
 Ion Ratio Lower Upper
 43 100
 58 28.2 0.0 61.4

