

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070325\
 Data File : VW031760.D
 Acq On : 03 Jul 2025 11:07
 Operator : SY/MD
 Sample : VW0703SBL02
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0703SBL02

Quant Time: Jul 04 04:50:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	208367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	400569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	360954	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	170260	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	140406	46.954	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	93.900%
35) Dibromofluoromethane	7.904	113	121852	46.617	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	93.240%
50) Toluene-d8	10.325	98	454815	46.756	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	93.520%
62) 4-Bromofluorobenzene	12.617	95	158138	44.177	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	88.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070325\
Data File : VW031760.D
Acq On : 03 Jul 2025 11:07
Operator : SY/MD
Sample : VW0703SBL02
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0703SBL02

Quant Time: Jul 04 04:50:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:35:17 2025
Response via : Initial Calibration

