

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070522\
 Data File : VW023562.D
 Acq On : 05 Jul 2022 10:10
 Operator : SY/VA
 Sample : VW0705SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK537

Quant Time: Jul 05 23:57:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM063022SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 05 23:57:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	354645	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	333129	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	156842	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	161453	31.827	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 127.320%		
7) Chloroethane-d5	2.879	69	80288	22.981	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 91.920%		
11) 1,1-Dichloroethene-d2	4.013	63	136643	21.770	ug/L	-0.01
Spiked Amount	25.000	Range 45 - 110	Recovery	= 87.080%		
21) 2-Butanone-d5	7.079	46	67036	52.761	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 105.520%		
24) Chloroform-d	7.646	84	256215	26.465	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 105.880%		
26) 1,2-Dichloroethane-d4	8.305	65	147024	27.501	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 110.000%		
32) Benzene-d6	8.274	84	498158	27.658	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 110.640%		
36) 1,2-Dichloropropane-d6	9.268	67	147136	27.631	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 110.520%		
41) Toluene-d8	10.323	98	481881	27.509	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 110.040%		
43) trans-1,3-Dichloroprop...	10.579	79	62388	26.447	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 105.800%		
47) 2-Hexanone-d5	10.920	63	51632	51.115	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 102.240%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	128647	25.162	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 100.640%		
66) 1,2-Dichlorobenzene-d4	13.853	152	157251	27.490	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 109.960%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070522\
Data File : VW023562.D
Acq On : 05 Jul 2022 10:10
Operator : SY/VA
Sample : VW0705SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK537

Quant Time: Jul 05 23:57:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM063022SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 05 23:57:26 2022
Response via : Initial Calibration

