

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW070618\
 Data File : VW003741.D
 Acq On : 06 Jul 2018 16:11
 Operator : SY/AP
 Sample : VSTDICV025
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
ClientSampleId :
 VICV86

Manual Integrations
APPROVED

MMDadoda
 7/9/2018 4:40:09 PM

Quant Time: Jul 07 00:15:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070618S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 06 15:27:20 2018
 Response via : Initial Calibration



TIC: VW003741.D

(2) Dichlorodifluoromethane (T)

2.008min (+0.006) 15.58ug/L

response 52197

Ion	Exp%	Act%
85.00	100	100
87.00	33.30	31.40
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW070618\
 Data File : VW003741.D
 Acq On : 06 Jul 2018 16:11
 Operator : SY/AP
 Sample : VSTDICV025
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VICV86

Manual Integrations
 APPROVED

MMDadoda
 7/9/2018 4:40:09 PM

Quant Time: Jul 07 00:15:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070618S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 06 15:27:20 2018
 Response via : Initial Calibration



TIC: VW003741.D

(2) Dichlorodifluoromethane (T)

2.008min (+0.006) 18.25ug/L m

response 61124

Ion	Exp%	Act%
85.00	100	100
87.00	33.30	26.81
0.00	0.00	0.00
0.00	0.00	0.00

Handwritten signature/initials

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW070618\
 Data File : VW003741.D
 Acq On : 06 Jul 2018 16:11
 Operator : SY/AP
 Sample : VSTDICV025
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VICV86

Manual Integrations
 APPROVED

MMDadoda
 7/9/2018 4:40:09 PM

Quant Time: Jul 07 01:17:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070618S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 06 15:27:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	380724	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	359745	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	186954	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	121816	23.39	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery =	93.56%		
7) Chloroethane-d5	2.89	69	87634	22.26	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery =	89.04%		
10) 1,1-Dichloroethene-d2	4.01	63	242742	25.98	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery =	103.92%		
20) 2-Butanone-d5	7.08	46	85690	52.02	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery =	104.04%		
24) Chloroform-d	7.65	84	240146	23.20	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery =	92.80%		
26) 1,2-Dichloroethane-d4	8.31	65	146671	23.56	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery =	94.24%		
29) Benzene-d6	8.27	84	462435	24.35	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery =	97.40%		
33) 1,2-Dichloropropane-d6	9.27	67	147964	24.21	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery =	96.84%		
37) Toluene-d8	10.32	98	448376	24.86	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery =	99.44%		
38) trans-1,3-Dichloropropene-	10.58	79	71325	25.20	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery =	100.80%		
39) 2-Hexanone-d5	10.93	63	65079	53.77	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery =	107.54%		
48) 1,1,2,2-Tetrachloroethane-	12.70	84	141832	24.29	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery =	97.16%		
61) 1,2-Dichlorobenzene-d4	13.87	152	167008	23.69	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery =	94.76%		

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	2.01	85	61124m	18.249	ug/L
3) Chloromethane	2.21	50	100541	22.499	ug/L
5) Vinyl chloride	2.36	62	135711	20.779	ug/L
6) Bromomethane	2.78	94	79804	21.001	ug/L
8) Chloroethane	2.93	64	75285	21.073	ug/L
9) Trichlorofluoromethane	3.25	101	64016	21.823	ug/L
11) 1,1,2-Trichloro-1,2,2-trif	4.06	101	116141	25.216	ug/L
12) 1,1-Dichloroethene	4.03	96	112019	24.889	ug/L
13) Acetone	4.12	43	83492	47.393	ug/L #
14) Carbon disulfide	4.37	76	371050	25.275	ug/L
15) Methyl Acetate	4.67	43	74421	26.511	ug/L
16) Methylene chloride	4.91	84	135197	22.192	ug/L
17) Methyl tert-butyl Ether	5.43	73	231400	25.827	ug/L
18) trans-1,2-Dichloroethene	5.42	96	122643	24.143	ug/L
19) 1,1-Dichloroethane	6.21	63	229577	23.177	ug/L
21) 2-Butanone	7.17	43	105556	45.455	ug/L
22) cis-1,2-Dichloroethene	7.17	96	133636	23.932	ug/L

07/07/18

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW070618\
 Data File : VW003741.D
 Acq On : 06 Jul 2018 16:11
 Operator : SY/AP
 Sample : VSTDICV025
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampled :
 VICV86

Manual Integrations
 APPROVED

MMDadoda
 7/9/2018 4:40:09 PM

Quant Time: Jul 07 01:17:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070618S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 06 15:27:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
23) Bromochloromethane	7.51	128	61878	23.424	ug/L	96
25) Chloroform	7.68	83	236199	22.941	ug/L	99
27) 1,2-Dichloroethane	8.40	62	177218	23.339	ug/L	100
30) Cyclohexane	7.96	56	209792	25.291	ug/L	99
31) 1,1,1-Trichloroethane	7.87	97	191380	24.137	ug/L	100
32) Carbon tetrachloride	8.07	117	186657	23.877	ug/L	100
34) Benzene	8.32	78	521766	24.145	ug/L	100
35) Trichloroethene	9.09	95	136237	23.798	ug/L	99
36) Methylcyclohexane	9.34	83	241855	25.138	ug/L	99
40) 1,2-Dichloropropane	9.37	63	139646	24.307	ug/L	99
41) Bromodichloromethane	9.65	83	177212	23.784	ug/L	98
42) cis-1,3-Dichloropropene	10.07	75	212906	25.019	ug/L	97
43) 4-Methyl-2-pentanone	10.21	43	215713	50.596	ug/L	99
44) Toluene	10.39	91	574719	24.888	ug/L	99
45) trans-1,3-Dichloropropene	10.61	75	197813	25.321	ug/L	97
46) 1,1,2-Trichloroethane	10.79	97	104813	23.779	ug/L	99
47) Tetrachloroethene	10.87	164	114767	23.999	ug/L	97
49) 2-Hexanone	10.98	43	167315	50.965	ug/L	99
50) Dibromochloromethane	11.13	129	126633	24.128	ug/L	99
51) 1,2-Dibromoethane	11.24	107	103250	24.265	ug/L	96
52) Chlorobenzene	11.66	112	367326	24.150	ug/L	96
53) Ethylbenzene	11.74	91	656462	25.217	ug/L	98
54) m,p-Xylene	11.85	106	244772	25.477	ug/L	99
55) o-xylene	12.17	106	234301	25.856	ug/L	95
56) Styrene	12.19	104	409711	25.736	ug/L	99
57) Isopropylbenzene	12.47	105	644790	26.002	ug/L	99
58) 1,1,2,2-Tetrachloroethane	12.72	83	143139	24.178	ug/L	99
59) 1,2,3-Trichloropropane	12.77	75	108513	24.434	ug/L	99
62) Bromoform	12.35	173	78384	22.975	ug/L	100
63) 1,3-Dichlorobenzene	13.51	146	284318	24.088	ug/L	98
64) 1,4-Dichlorobenzene	13.59	146	300658	23.412	ug/L	98
65) 1,2-Dichlorobenzene	13.88	146	276391	23.859	ug/L	96
66) 1,2-Dibromo-3-chloropropan	14.49	75	25154	23.254	ug/L	96
67) 1,3,5-Trichlorobenzene	14.64	180	225208	24.174	ug/L	99
68) 1,2,4-trichlorobenzene	15.15	180	183778	24.445	ug/L	99
69) Naphthalene	15.38	128	393027	25.663	ug/L	99
70) 1,2,3-Trichlorobenzene	15.57	180	174159	24.306	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW070618\
Data File : VW003741.D
Acq On : 06 Jul 2018 16:11
Operator : SY/AP
Sample : VSTDICV025
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
Client Sampled :
VICV86

Manual Integrations
APPROVED

MMDadoda
7/9/2018 4:40:09 PM

Quant Time: Jul 07 01:17:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070618S.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 06 15:27:20 2018
Response via : Initial Calibration

