

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070623\
 Data File : VW026541.D
 Acq On : 06 Jul 2023 14:29
 Operator : SY/MD
 Sample : 03444-10
 Misc : 5.02g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jul 07 02:34:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 04 13:17:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	855899	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	772934	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	393653	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	260697	20.256	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.040%
7) Chloroethane-d5	2.893	69	200221	18.804	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.200%
11) 1,1-Dichloroethene-d2	4.027	65	128408	21.391	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.560%
21) 2-Butanone-d5	7.075	46	203649	41.363	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.720%
24) Chloroform-d	7.648	84	529551	21.001	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.000%
26) 1,2-Dichloroethane-d4	8.307	65	310619	21.179	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.720%
32) Benzene-d6	8.276	84	1140645	23.146	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	9.276	67	358255	22.801	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.200%
41) Toluene-d8	10.318	98	1002039	22.649	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.600%
43) trans-1,3-Dichloroprop...	10.575	79	139837	21.944	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.760%
47) 2-Hexanone-d5	10.922	63	128488	41.072	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.140%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	292906	21.305	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.240%
66) 1,2-Dichlorobenzene-d4	13.854	152	300586	19.226	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070623\
Data File : VW026541.D
Acq On : 06 Jul 2023 14:29
Operator : SY/MD
Sample : 03444-10
Misc : 5.02g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jul 07 02:34:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070323SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 04 13:17:14 2023
Response via : Initial Calibration

