

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070822\
 Data File : VW023589.D
 Acq On : 08 Jul 2022 12:08
 Operator : SY/VA
 Sample : N3598-11
 Misc : 6.17g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AG4

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 07/09/2022
 Supervised By : Mahesh Dadoda 07/11/2022

Quant Time: Jul 08 23:21:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 08 23:21:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	314249	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	300082	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	142287	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	108041	18.993	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	75.960%		
7) Chloroethane-d5	2.879	69	65886	20.252	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	81.000%		
11) 1,1-Dichloroethene-d2	4.013	63	98944	15.428	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	61.720%		
21) 2-Butanone-d5	7.074	46	64290	54.424	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	108.840%		
24) Chloroform-d	7.647	84	203366	22.720	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	90.880%		
26) 1,2-Dichloroethane-d4	8.299	65	118638m	23.450	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	93.800%		
32) Benzene-d6	8.269	84	374979	22.584	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	90.320%		
36) 1,2-Dichloropropane-d6	9.274	67	116925	24.053	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	96.200%		
41) Toluene-d8	10.323	98	352286	21.890	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	87.560%		
43) trans-1,3-Dichloroprop...	10.573	79	45840	21.126	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	84.520%		
47) 2-Hexanone-d5	10.921	63	47873	50.354	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	100.700%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	116230	24.101	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	96.400%		
66) 1,2-Dichlorobenzene-d4	13.853	152	120438	23.461	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	93.840%		
Target Compounds					Qvalue	
13) Acetone	4.129	43	3777m	4.338	ug/L	
34) Trichloroethene	9.092	95	4991	1.016	ug/L	87
46) Tetrachloroethene	10.866	164	19362	5.016	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070822\
Data File : VW023589.D
Acq On : 08 Jul 2022 12:08
Operator : SY/VA
Sample : N3598-11
Misc : 6.17g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AG4

Quant Time: Jul 08 23:21:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 08 23:21:00 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/09/2022
Supervised By :Mahesh Dadoda 07/11/2022

