

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070924\
 Data File : VW029482.D
 Acq On : 09 Jul 2024 10:53
 Operator : SY/MD
 Sample : VW0709SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK517

Quant Time: Jul 10 01:19:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 10 01:18:26 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.843	114	268843	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.629	117	256798	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.556	152	115021	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.363	65	109945	25.316	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.280%	
7) Chloroethane-d5		2.905	69	83488	26.698	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.800%	
11) 1,1-Dichloroethene-d2		4.021	65	50052	25.665	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	102.640%	
21) 2-Butanone-d5		7.075	46	59688	45.783	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.560%	
24) Chloroform-d		7.648	84	210692	27.109	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	108.440%	
26) 1,2-Dichloroethane-d4		8.307	65	124631	27.339	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	109.360%	
32) Benzene-d6		8.276	84	420344	27.324	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	109.280%	
36) 1,2-Dichloropropane-d6		9.270	67	133184	27.077	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.320%	
41) Toluene-d8		10.319	98	374887	27.710	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	110.840%	
43) trans-1,3-Dichloroprop...		10.575	79	51711	24.827	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.320%	
47) 2-Hexanone-d5		10.922	63	40729	45.569	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.140%	
56) 1,1,2,2-Tetrachloroeth...		12.690	84	98792	24.738	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.960%	
66) 1,2-Dichlorobenzene-d4		13.849	152	114142	28.495	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	114.000%	
Target Compounds							Qvalue
16) Methylene chloride		4.923	84	14945	3.105	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070924\
Data File : VW029482.D
Acq On : 09 Jul 2024 10:53
Operator : SY/MD
Sample : VW0709SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK517

Quant Time: Jul 10 01:19:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 10 01:18:26 2024
Response via : Initial Calibration

