

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070924\
 Data File : VW029486.D
 Acq On : 09 Jul 2024 16:10
 Operator : SY/MD
 Sample : P3177-04
 Misc : 5.05g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BS2

Quant Time: Jul 10 01:20:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 10 01:18:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	391350	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	329107	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	100400	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	107085	16.938	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.760%
7) Chloroethane-d5	2.893	69	79359	17.433	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.720%
11) 1,1-Dichloroethene-d2	4.021	65	49361	17.387	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.560%
21) 2-Butanone-d5	7.081	46	47247	24.896	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	49.800%
24) Chloroform-d	7.648	84	203920	18.024	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.080%
26) 1,2-Dichloroethane-d4	8.307	65	111600	16.817	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	67.280%#
32) Benzene-d6	8.276	84	399032	20.240	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.960%
36) 1,2-Dichloropropane-d6	9.270	67	125597	19.924	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.680%
41) Toluene-d8	10.319	98	346252	19.970	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.880%
43) trans-1,3-Dichloroprop...	10.575	79	45438	17.022	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.080%
47) 2-Hexanone-d5	10.922	63	31290	27.317	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.640%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	73219	14.306	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	57.240%
66) 1,2-Dichlorobenzene-d4	13.849	152	63280	18.098	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.400%#
Target Compounds					Qvalue	
13) Acetone	4.125	43	13993	7.305	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070924\
Data File : VW029486.D
Acq On : 09 Jul 2024 16:10
Operator : SY/MD
Sample : P3177-04
Misc : 5.05g/10mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BS2

Quant Time: Jul 10 01:20:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 10 01:18:26 2024
Response via : Initial Calibration

