

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070925\
 Data File : VW031775.D
 Acq On : 09 Jul 2025 14:56
 Operator : SY/MD
 Sample : Q2514-06
 Misc : 5.10g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-103

Quant Time: Jul 10 01:26:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	196576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	400527	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	357765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	165193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	140683	49.868	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	99.740%
35) Dibromofluoromethane	7.904	113	125086	47.859	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	95.720%
50) Toluene-d8	10.331	98	460855	47.382	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	94.760%
62) 4-Bromofluorobenzene	12.617	95	165023	46.105	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	92.200%
Target Compounds						
20) Methylene Chloride	4.954	84	36129	9.246	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070925\
Data File : VW031775.D
Acq On : 09 Jul 2025 14:56
Operator : SY/MD
Sample : Q2514-06
Misc : 5.10g/5mL/MSVOA_W/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-103

Quant Time: Jul 10 01:26:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:35:17 2025
Response via : Initial Calibration

