

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071024\
 Data File : VW029503.D
 Acq On : 10 Jul 2024 13:50
 Operator : SY/MD
 Sample : P3177-09RE
 Misc : 5.40g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BS7RE

Quant Time: Jul 11 02:16:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jul 11 02:12:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	272123	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	157879	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	32369	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	115301	26.229	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.920%
7) Chloroethane-d5	2.893	69	87608	27.678	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.720%
11) 1,1-Dichloroethene-d2	4.021	65	48803	24.723	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	98.880%
21) 2-Butanone-d5	7.075	46	36191	27.426	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.860%
24) Chloroform-d	7.648	84	204162	25.952	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.800%
26) 1,2-Dichloroethane-d4	8.307	65	126989	27.521	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	110.080%
32) Benzene-d6	8.276	84	387810	41.004	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	164.000%#
36) 1,2-Dichloropropane-d6	9.270	67	128541	42.507	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	170.040%#
41) Toluene-d8	10.325	98	287537	34.569	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	138.280%#
43) trans-1,3-Dichloroprop...	10.575	79	39298	30.688	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	122.760%
47) 2-Hexanone-d5	10.922	63	26224	47.723	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.440%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	72627	29.581	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	118.320%
66) 1,2-Dichlorobenzene-d4	13.848	152	40144	35.612	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	142.440%#
Target Compounds					Qvalue	
16) Methylene chloride	4.929	84	10711	2.199	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071024\
Data File : VW029503.D
Acq On : 10 Jul 2024 13:50
Operator : SY/MD
Sample : P3177-09RE
Misc : 5.40g/10mL/MSVOA_W/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BS7RE

Quant Time: Jul 11 02:16:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jul 11 02:12:03 2024
Response via : Initial Calibration

