

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071025\
 Data File : VW031804.D
 Acq On : 10 Jul 2025 15:36
 Operator : SY/MD
 Sample : Q2555-01
 Misc : 4.35g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OU4-TS-29-070925

Quant Time: Jul 11 02:22:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	166057	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.849	114	358196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	316208	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	137883	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	128960	54.114	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	108.220%
35) Dibromofluoromethane	7.898	113	111242	47.592	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	95.180%
50) Toluene-d8	10.331	98	393680	45.259	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	90.520%
62) 4-Bromofluorobenzene	12.617	95	138843	43.375	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	86.740%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071025\
Data File : VW031804.D
Acq On : 10 Jul 2025 15:36
Operator : SY/MD
Sample : Q2555-01
Misc : 4.35g/5mL/MSVOA_W/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OU4-TS-29-070925

Quant Time: Jul 11 02:22:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:35:17 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.965 min Scan# 997
Delta R.T. 0.000 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36




Instrument :
MSVOA_W
ClientSampleId :
OU4-TS-29-070925


Tgt Ion:168 Resp: 166057
Ion Ratio Lower Upper
168 100
99 75.7 49.4 74.2#



#33
1,2-Dichloroethane-d4
Concen: 54.114 ug/l
RT: 8.319 min Scan# 1055
Delta R.T. 0.000 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36

Tgt Ion: 65 Resp: 128960
Ion Ratio Lower Upper
65 100
67 51.0 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.849 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36




Instrument :
MSVOA_W
ClientSampleId :
OU4-TS-29-070925


Tgt Ion:114 Resp: 358196
Ion Ratio Lower Upper
114 100
63 21.7 0.0 43.6
88 17.2 0.0 34.2



#35
Dibromofluoromethane
Concen: 47.592 ug/l
RT: 7.898 min Scan# 986
Delta R.T. -0.006 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36

Tgt Ion:113 Resp: 111242
Ion Ratio Lower Upper
113 100
111 102.2 82.1 123.1
192 16.1 13.8 20.6









#50
Toluene-d8
Concen: 45.259 ug/l
RT: 10.331 min Scan# 11
Delta R.T. 0.006 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36




Instrument :
MSVOA_W
ClientSampleId :
OU4-TS-29-070925


Tgt Ion: 98 Resp: 393680
Ion Ratio Lower Upper
98 100
100 64.7 53.0 79.4



#62
4-Bromofluorobenzene
Concen: 43.375 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. 0.000 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36

Tgt Ion: 95 Resp: 138843
Ion Ratio Lower Upper
95 100
174 65.8 0.0 133.8
176 60.1 0.0 126.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.629 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36

Instrument :
MSVOA_W
ClientSampleId :
OU4-TS-29-070925

Tgt Ion:117 Resp: 316208
Ion Ratio Lower Upper
117 100
82 61.2 48.6 72.8
119 31.7 23.9 35.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.556 min Scan# 1914
Delta R.T. 0.000 min
Lab File: VW031804.D
Acq: 10 Jul 2025 15:36

Tgt Ion:152 Resp: 137883
Ion Ratio Lower Upper
152 100
115 65.7 31.9 95.7
150 157.3 0.0 356.4

