

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071123\
 Data File : VW026593.D
 Acq On : 11 Jul 2023 13:46
 Operator : SY/MD
 Sample : 03551-08
 Misc : 6.42g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCHB5

Quant Time: Jul 12 07:16:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 12 07:03:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	862068	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	740785	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	333658	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	277579	19.089	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.360%
7) Chloroethane-d5	2.893	69	200094	19.610	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.440%
11) 1,1-Dichloroethene-d2	4.027	65	134037	19.628	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.520%
21) 2-Butanone-d5	7.075	46	176454	37.547	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.100%
24) Chloroform-d	7.648	84	495560	19.967	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.880%
26) 1,2-Dichloroethane-d4	8.307	65	280731	20.258	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.040%
32) Benzene-d6	8.276	84	1074329	21.126	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.520%
36) 1,2-Dichloropropane-d6	9.276	67	329812	21.092	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.360%
41) Toluene-d8	10.324	98	923137	21.165	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.640%
43) trans-1,3-Dichloroprop...	10.581	79	123294	19.676	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.720%
47) 2-Hexanone-d5	10.922	63	113121	37.075	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.140%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	255540	19.298	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	77.200%
66) 1,2-Dichlorobenzene-d4	13.854	152	261506	21.347	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071123\
Data File : VW026593.D
Acq On : 11 Jul 2023 13:46
Operator : SY/MD
Sample : 03551-08
Misc : 6.42g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCHB5

Quant Time: Jul 12 07:16:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 12 07:03:59 2023
Response via : Initial Calibration

