

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071123\
 Data File : VW026590.D
 Acq On : 11 Jul 2023 12:13
 Operator : SY/MD
 Sample : VW0711SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK536

Quant Time: Jul 12 07:16:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 12 07:03:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	659842	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	581853	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	256339	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	215920	19.399	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.600%
7) Chloroethane-d5	2.893	69	157605	20.180	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.720%
11) 1,1-Dichloroethene-d2	4.021	65	104007	19.898	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.600%
21) 2-Butanone-d5	7.075	46	166996	46.425	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.840%
24) Chloroform-d	7.648	84	398610	20.983	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.920%
26) 1,2-Dichloroethane-d4	8.307	65	232871	21.954	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.800%
32) Benzene-d6	8.276	84	850955	21.304	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.200%
36) 1,2-Dichloropropane-d6	9.270	67	262633	21.384	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.520%
41) Toluene-d8	10.325	98	730152	21.313	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.240%
43) trans-1,3-Dichloroprop...	10.575	79	101128	20.547	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.200%
47) 2-Hexanone-d5	10.922	63	113500	47.360	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	226273	21.756	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.040%
66) 1,2-Dichlorobenzene-d4	13.854	152	224633	23.868	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071123\
Data File : VW026590.D
Acq On : 11 Jul 2023 12:13
Operator : SY/MD
Sample : VW0711SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK536

Quant Time: Jul 12 07:16:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 12 07:03:59 2023
Response via : Initial Calibration

