

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071124\
 Data File : VW029518.D
 Acq On : 11 Jul 2024 10:37
 Operator : SY/MD
 Sample : VW0711SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK519

Quant Time: Jul 12 04:32:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 12 04:31:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	299029	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	277812	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	120836	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	88584	18.338	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	73.360%		
7) Chloroethane-d5	2.899	69	67435	19.387	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	77.560%		
11) 1,1-Dichloroethene-d2	4.021	65	40901	18.855	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	75.440%		
21) 2-Butanone-d5	7.075	46	44332	30.572	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	61.140%		
24) Chloroform-d	7.648	84	173212	20.037	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.160%		
26) 1,2-Dichloroethane-d4	8.307	65	101450	20.008	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	80.040%		
32) Benzene-d6	8.276	84	339645	20.408	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	81.640%		
36) 1,2-Dichloropropane-d6	9.270	67	110386	20.745	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	82.960%		
41) Toluene-d8	10.325	98	298189	20.373	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	81.480%		
43) trans-1,3-Dichloroprop...	10.575	79	41279	18.319	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	73.280%		
47) 2-Hexanone-d5	10.922	63	29892	30.914	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	61.820%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	76791	17.774	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	71.080%		
66) 1,2-Dichlorobenzene-d4	13.848	152	96226	22.866	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	91.480%		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	25591	4.780	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071124\
Data File : VW029518.D
Acq On : 11 Jul 2024 10:37
Operator : SY/MD
Sample : VW0711SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK519

Quant Time: Jul 12 04:32:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 12 04:31:40 2024
Response via : Initial Calibration

