

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071124\
 Data File : VW029529.D
 Acq On : 11 Jul 2024 15:10
 Operator : SY/MD
 Sample : P3178-09
 Misc : 4.88g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BT4

Quant Time: Jul 12 04:34:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 12 04:31:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	337189	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	224389	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	57604	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	103995	19.092	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	76.360%		
7) Chloroethane-d5	2.899	69	81305	20.730	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	82.920%		
11) 1,1-Dichloroethene-d2	4.021	65	45008	18.401	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	73.600%		
21) 2-Butanone-d5	7.075	46	45438	27.789	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	55.580%		
24) Chloroform-d	7.648	84	182319	18.703	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	74.800%		
26) 1,2-Dichloroethane-d4	8.307	65	111847	19.562	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	78.240%		
32) Benzene-d6	8.276	84	346040	25.743	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	102.960%		
36) 1,2-Dichloropropane-d6	9.276	67	115310	26.829	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	107.320%		
41) Toluene-d8	10.325	98	244885	20.715	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	82.840%		
43) trans-1,3-Dichloroprop...	10.575	79	31986	17.575	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	70.280%		
47) 2-Hexanone-d5	10.922	63	29425	37.677	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	75.360%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	60382	17.304	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	69.200%		
66) 1,2-Dichlorobenzene-d4	13.849	152	32962	16.431	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	65.720%#		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	26247	4.348	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071124\
Data File : VW029529.D
Acq On : 11 Jul 2024 15:10
Operator : SY/MD
Sample : P3178-09
Misc : 4.88g/10mL/MSVOA_W/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BT4

Quant Time: Jul 12 04:34:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 12 04:31:40 2024
Response via : Initial Calibration

