

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071125\
 Data File : VW031834.D
 Acq On : 11 Jul 2025 20:56
 Operator : SY/MD
 Sample : VSTDCCC050EC
 Misc : 5.00g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jul 11 22:54:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	1300	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.641	117	2012	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.556	152	1520	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.337	65	412	39.219	ug/l	0.02
Spiked Amount	50.000	Range	63 - 155	Recovery	=	78.440%
35) Dibromofluoromethane	7.904	113	181	21.337	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	42.680%#
50) Toluene-d8	10.325	98	1233	39.057	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	78.120%
62) 4-Bromofluorobenzene	12.617	95	1597	137.466	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	274.940%#

Target Compounds			Qvalue			
2) Dichlorodifluoromethane	2.046	85	44	8.826	ug/l #	44
3) Chloromethane	2.259	50	339	56.405	ug/l #	67
4) Vinyl Chloride	2.491	62	46	5.823	ug/l #	41
5) Bromomethane	2.826	94	139	22.527	ug/l	88
6) Chloroethane	3.003	64	22	4.100	ug/l #	44
8) Diethyl Ether	3.728	74	1782	326.883	ug/l	61
11) Tert butyl alcohol	5.198	59	18	27.658	ug/l #	69
12) 1,1-Dichloroethene	4.027	96	43	4.930	ug/l #	1
13) Acrolein	3.936	56	103	88.054	ug/l #	1
14) Allyl chloride	4.704	41	134	10.013	ug/l #	20
15) Acrylonitrile	5.429	53	353	131.677	ug/l #	46
16) Acetone	4.167	43	3413	1314.451	ug/l #	81
17) Carbon Disulfide	4.423	76	224	9.597	ug/l #	77
18) Methyl Acetate	4.704	43	49	6.603	ug/l #	49
20) Methylene Chloride	4.960	84	27348	Below Cal		94
21) trans-1,2-Dichloroethene	5.454	96	45	4.855	ug/l #	1
23) Vinyl Acetate	6.301	43	285	15.987	ug/l #	70
24) 1,1-Dichloroethane	6.356	63	24	1.418	ug/l #	82
25) 2-Butanone	7.167	43	37	10.953	ug/l #	46
26) 2,2-Dichloropropane	7.179	77	22	2.228	ug/l #	52
27) cis-1,2-Dichloroethene	7.173	96	20	1.877	ug/l #	1
29) Tetrahydrofuran	7.575	42	291	125.792	ug/l #	48
30) Chloroform	7.697	83	457	25.417	ug/l	91
32) 1,1,1-Trichloroethane	7.880	97	22	1.588	ug/l #	24
36) 1,1-Dichloropropene	8.106	75	23	1.874	ug/l #	42
37) Ethyl Acetate	7.282	43	143	18.467	ug/l #	68
40) Benzene	8.343	78	491	13.519	ug/l #	51
41) Methacrylonitrile	7.478	41	19	4.580	ug/l #	1
42) 1,2-Dichloroethane	8.404	62	27	2.200	ug/l #	76
43) Isopropyl Acetate	8.410	43	21	1.573	ug/l #	52
46) Dibromomethane	9.483	93	38	6.706	ug/l #	33
47) Bromodichloromethane	9.642	83	75	5.644	ug/l #	25
48) Methyl methacrylate	9.459	41	72	10.970	ug/l #	1
51) 4-Methyl-2-Pentanone	10.209	43	286	37.268	ug/l #	48
52) Toluene	10.398	92	379	16.299	ug/l	90
53) t-1,3-Dichloropropene	10.617	75	166	13.314	ug/l #	1

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071125\
 Data File : VW031834.D
 Acq On : 11 Jul 2025 20:56
 Operator : SY/MD
 Sample : VSTDCCC050EC
 Misc : 5.00g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1


Instrument :
 MSVOA_W
 ClientSampleId :


Quant Time: Jul 11 22:54:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
54) cis-1,3-Dichloropropene	10.087	75	142	10.046	ug/l #	67
55) 1,1,2-Trichloroethane	10.770	97	28	3.772	ug/l #	17
56) Ethyl methacrylate	10.678	69	26	2.582	ug/l #	1
57) 1,3-Dichloropropane	10.934	76	33	2.532	ug/l #	42
58) 2-Chloroethyl Vinyl ether	10.087	63	19	3.426	ug/l #	45
59) 2-Hexanone	10.983	43	145	27.883	ug/l #	23
61) 1,2-Dibromoethane	11.245	107	114	15.341	ug/l #	52
65) Chlorobenzene	11.666	112	382	8.605	ug/l #	45
67) Ethyl Benzene	11.733	91	396	5.144	ug/l #	43
68) m/p-Xylenes	11.855	106	267	9.023	ug/l #	1
69) o-Xylene	12.178	106	105	3.832	ug/l #	1
70) Styrene	12.196	104	542	11.443	ug/l #	80
73) Isopropylbenzene	12.458	105	239	2.067	ug/l #	49
74) N-amyl acetate	12.294	43	473	15.178	ug/l #	60
75) 1,1,2,2-Tetrachloroethane	12.714	83	190	7.402	ug/l #	38
76) 1,2,3-Trichloropropane	12.757	75	60	3.055	ug/l #	1
77) Bromobenzene	12.739	156	125	4.847	ug/l #	32
78) n-propylbenzene	12.800	91	569	4.110	ug/l #	74
79) 2-Chlorotoluene	13.001	91	927	11.119	ug/l	73
80) 1,3,5-Trimethylbenzene	12.940	105	461	4.817	ug/l	89
81) trans-1,4-Dichloro-2-b...	12.623	75	815	96.640	ug/l #	8
82) 4-Chlorotoluene	13.001	91	927	10.591	ug/l	73
83) tert-Butylbenzene	13.196	119	186	2.269	ug/l #	16
84) 1,2,4-Trimethylbenzene	13.245	105	886	9.137	ug/l	95
85) sec-Butylbenzene	13.385	105	507	4.120	ug/l #	66
86) p-Isopropyltoluene	13.495	119	494	4.871	ug/l #	49
87) 1,3-Dichlorobenzene	13.580	146	407	7.870	ug/l	81
88) 1,4-Dichlorobenzene	13.580	146	407	7.689	ug/l	75
89) n-Butylbenzene	13.818	91	508	5.156	ug/l #	87
90) Hexachloroethane	14.104	117	46	2.514	ug/l #	19
91) 1,2-Dichlorobenzene	13.861	146	385	8.145	ug/l #	67
93) 1,2,4-Trichlorobenzene	15.123	180	297	10.189	ug/l #	53
95) Naphthalene	15.366	128	2236	31.288	ug/l #	88
96) 1,2,3-Trichlorobenzene	15.555	180	416	15.531	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jul 11 22:54:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:35:17 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.959 min Scan# 99
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Instrument :
MSVOA_W
ClientSampleId :







Tgt Ion:168 Resp: 732
Ion Ratio Lower Upper
168 100
99 58.7 49.4 74.2



#2
Dichlorodifluoromethane
Concen: 8.826 ug/l
RT: 2.046 min Scan# 26
Delta R.T. 0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 85 Resp: 44
Ion Ratio Lower Upper
85 100
87 0.0 15.4 46.2#









#3
Chloromethane
Concen: 56.405 ug/l
RT: 2.259 min Scan# 61
Delta R.T. 0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 50 Resp: 339
Ion Ratio Lower Upper
50 100
52 51.4 26.2 39.2#



#4
Vinyl Chloride
Concen: 5.823 ug/l
RT: 2.491 min Scan# 99
Delta R.T. 0.086 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 62 Resp: 46
Ion Ratio Lower Upper
62 100
64 0.0 26.8 40.2#









#5
Bromomethane
Concen: 22.527 ug/l
RT: 2.826 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 94 Resp: 139
Ion Ratio Lower Upper
94 100
96 106.1 75.6 113.4



#6
Chloroethane
Concen: 4.100 ug/l
RT: 3.003 min Scan# 183
Delta R.T. 0.031 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 64 Resp: 22
Ion Ratio Lower Upper
64 100
66 0.0 24.5 36.7#





#8

Diethyl Ether

Concen: 326.883 ug/l

RT: 3.728 min Scan# 301

Delta R.T. 0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 74 Resp: 1782

Ion Ratio Lower Upper

74 100

45 124.4 43.9 131.7



#11

Tert butyl alcohol

Concen: 27.658 ug/l

RT: 5.198 min Scan# 543

Delta R.T. -0.018 min

Lab File: VW031834.D




Acq: 11 Jul 2025 20:56


Tgt Ion: 59 Resp: 18

Ion Ratio Lower Upper

59 100

57 0.0 9.5 14.3#









#12
1,1-Dichloroethene
Concen: 4.930 ug/l
RT: 4.027 min Scan# 351
Delta R.T. -0.048 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 96 Resp: 43
Ion Ratio Lower Upper
96 100
61 0.0 135.8 203.8#
98 0.0 50.4 75.6#



#13
Acrolein
Concen: 88.054 ug/l
RT: 3.936 min Scan# 336
Delta R.T. 0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56


Tgt Ion: 56 Resp: 103
Ion Ratio Lower Upper
56 100
55 158.3 54.0 81.0#








#14
 Allyl chloride
 Concen: 10.013 ug/l
 RT: 4.704 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VW031834.D
 Acq: 11 Jul 2025 20:56

Instrument :
 MSVOA_W
 ClientSampleId :







Tgt Ion: 41 Resp: 134
 Ion Ratio Lower Upper
 41 100
 39 0.0 60.7 91.1#
 76 0.0 30.3 45.5#



#15
 Acrylonitrile
 Concen: 131.677 ug/l
 RT: 5.429 min Scan# 581
 Delta R.T. 0.031 min
 Lab File: VW031834.D
 Acq: 11 Jul 2025 20:56

Tgt Ion: 53 Resp: 353
 Ion Ratio Lower Upper
 53 100
 52 33.4 66.4 99.6#
 51 6.5 29.9 44.9#









#16
Acetone
Concen: 1314.451 ug/l
RT: 4.167 min Scan# 374
Delta R.T. 0.012 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 43 Resp: 3413
Ion Ratio Lower Upper
43 100
58 20.6 24.8 37.2#



#17
Carbon Disulfide
Concen: 9.597 ug/l
RT: 4.423 min Scan# 416
Delta R.T. 0.000 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 76 Resp: 224
Ion Ratio Lower Upper
76 100
78 0.0 6.6 10.0#









#18
Methyl Acetate
Concen: 6.603 ug/l
RT: 4.704 min Scan# 4
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 43 Resp: 49
Ion Ratio Lower Upper
43 100
74 0.0 20.6 30.8#



#20
Methylene Chloride
Concen: Below Cal
RT: 4.960 min Scan# 504
Delta R.T. 0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 84 Resp: 27348
Ion Ratio Lower Upper
84 100
49 111.2 98.2 147.4
51 36.5 29.8 44.8
86 60.6 49.5 74.3









#21
trans-1,2-Dichloroethene
Concen: 4.855 ug/l
RT: 5.454 min Scan# 51
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 96 Resp: 45
Ion Ratio Lower Upper
96 100
61 0.0 109.8 164.6#
98 0.0 54.9 82.3#



#23
Vinyl Acetate
Concen: 15.987 ug/l
RT: 6.301 min Scan# 724
Delta R.T. 0.019 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56


Tgt Ion: 43 Resp: 285
Ion Ratio Lower Upper
43 100
86 0.0 9.2 13.8#








#24
1,1-Dichloroethane
Concen: 1.418 ug/l
RT: 6.356 min Scan# 715
Delta R.T. 0.110 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Instrument :
MSVOA_W
ClientSampleId :







Tgt Ion: 63 Resp: 24
Ion Ratio Lower Upper
63 100
98 0.0 3.5 10.6#
100 0.0 2.5 7.5#



#25
2-Butanone
Concen: 10.953 ug/l
RT: 7.167 min Scan# 866
Delta R.T. -0.024 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 43 Resp: 37
Ion Ratio Lower Upper
43 100
72 0.0 23.0 34.6#





#26

2,2-Dichloropropane

Concen: 2.228 ug/l

RT: 7.179 min Scan# 8

Delta R.T. -0.012 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 77 Resp: 22

Ion Ratio Lower Upper

77 100

97 0.0 11.6 34.7#



#27

cis-1,2-Dichloroethene

Concen: 1.877 ug/l

RT: 7.173 min Scan# 867

Delta R.T. -0.018 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56




Tgt Ion: 96 Resp: 20


Ion Ratio Lower Upper

96 100

61 0.0 0.0 286.8

98 0.0 0.0 130.4





#29

Tetrahydrofuran

Concen: 125.792 ug/l

RT: 7.575 min Scan# 91

Delta R.T. 0.025 min


Lab File: VW031834.D

Acq: 11 Jul 2025 20:56




Instrument :

MSVOA_W

ClientSampleId :



Tgt Ion:	42	Resp:	291
Ion	Ratio	Lower	Upper
42	100		
72	0.0	36.8	55.2#
71	21.3	34.6	52.0#



#30

Chloroform

Concen: 25.417 ug/l




RT: 7.697 min Scan# 953


Delta R.T. 0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Tgt Ion:	83	Resp:	457
Ion	Ratio	Lower	Upper
83	100		
85	56.9	51.4	77.0









#32
 1,1,1-Trichloroethane
 Concen: 1.588 ug/l
 RT: 7.880 min Scan# 91
 Delta R.T. -0.012 min
 Lab File: VW031834.D
 Acq: 11 Jul 2025 20:56




Instrument :
 MSVOA_W
 ClientSampleId :


Tgt Ion: 97 Resp: 22
 Ion Ratio Lower Upper
 97 100
 99 0.0 51.0 76.6#
 61 0.0 36.0 54.0#



#33
 1,2-Dichloroethane-d4
 Concen: 39.219 ug/l
 RT: 8.337 min Scan# 1058
 Delta R.T. 0.019 min
 Lab File: VW031834.D
 Acq: 11 Jul 2025 20:56

Tgt Ion: 65 Resp: 412
 Ion Ratio Lower Upper
 65 100
 67 35.2 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.849 min Scan# 1142

Delta R.T. 0.000 min

Lab File: VW031834.D





Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :

Tgt Ion:114	Resp:	1300
Ion Ratio	Lower	Upper
114	100	
63	16.7	0.0 43.6
88	17.1	0.0 34.2



#35

Dibromofluoromethane

Concen: 21.337 ug/l




RT: 7.904 min Scan# 987


Delta R.T. 0.000 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Tgt Ion:113	Resp:	181
Ion Ratio	Lower	Upper
113	100	
111	121.0	82.1 123.1
192	0.0	13.8 20.6









#36
1,1-Dichloropropene
Concen: 1.874 ug/l
RT: 8.106 min Scan# 1018
Delta R.T. 0.012 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 75 Resp: 23
Ion Ratio Lower Upper
75 100
110 0.0 16.8 50.4#
77 0.0 25.0 37.6#



#37
Ethyl Acetate
Concen: 18.467 ug/l
RT: 7.282 min Scan# 885
Delta R.T. 0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 43 Resp: 143
Ion Ratio Lower Upper
43 100
61 0.0 10.9 16.3#
70 0.0 9.2 13.8#





#40

Benzene

Concen: 13.519 ug/l

RT: 8.343 min Scan# 1059

Delta R.T. 0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 78 Resp: 491

Ion Ratio Lower Upper

78 100

77 0.0 19.2 28.8#



#41

Methacrylonitrile

Concen: 4.580 ug/l

RT: 7.478 min Scan# 917

Delta R.T. -0.018 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Tgt Ion: 41 Resp: 19




Ion Ratio Lower Upper


41 100

39 163.2 43.2 64.8#

67 0.0 58.1 87.1#

52 0.0 25.6 38.4#





#42

1,2-Dichloroethane

Concen: 2.200 ug/l

RT: 8.404 min Scan# 1069

Delta R.T. -0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 62 Resp: 27

Ion Ratio Lower Upper

62 100

98 0.0 0.0 17.2



#43

Isopropyl Acetate

Concen: 1.573 ug/l

RT: 8.410 min Scan# 1070

Delta R.T. -0.024 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56




Tgt Ion: 43 Resp: 21


Ion Ratio Lower Upper

43 100

61 0.0 23.1 34.7#

87 0.0 10.8 16.2#





#46

Dibromomethane

Concen: 6.706 ug/l

RT: 9.483 min Scan# 11

Delta R.T. 0.012 min


Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :






Tgt Ion: 93 Resp: 38

Ion Ratio Lower Upper

93 100

95 50.0 67.0 100.4#

174 0.0 71.7 107.5#



#47

Bromodichloromethane

Concen: 5.644 ug/l

RT: 9.642 min Scan# 1272

Delta R.T. -0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56




Tgt Ion: 83 Resp: 75


Ion Ratio Lower Upper

83 100

85 0.0 50.8 76.2#


127 0.0 6.8 10.2#








#48
Methyl methacrylate
Concen: 10.970 ug/l
RT: 9.459 min Scan# 11
Delta R.T. 0.019 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Instrument :
MSVOA_W
ClientSampleId :







Tgt Ion: 41 Resp: 72
Ion Ratio Lower Upper
41 100
69 0.0 71.6 107.4#
39 170.8 49.5 74.3#



#50
Toluene-d8
Concen: 39.057 ug/l
RT: 10.325 min Scan# 1384
Delta R.T. 0.000 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 98 Resp: 1233
Ion Ratio Lower Upper
98 100
100 68.8 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 37.268 ug/l
RT: 10.209 min Scan# 1365
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 43 Resp: 286
Ion Ratio Lower Upper
43 100
58 8.0 32.2 48.2#



#52
Toluene
Concen: 16.299 ug/l
RT: 10.398 min Scan# 1396
Delta R.T. 0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 92 Resp: 379
Ion Ratio Lower Upper
92 100
91 160.7 139.5 209.3









#53
 t-1,3-Dichloropropene
 Concen: 13.314 ug/l
 RT: 10.617 min Scan# 1431
 Delta R.T. 0.006 min
 Lab File: VW031834.D
 Acq: 11 Jul 2025 20:56




Instrument :
 MSVOA_W
 ClientSampleId :


Tgt Ion: 75 Resp: 166
 Ion Ratio Lower Upper
 75 100
 77 94.0 24.3 36.5#



#54
 cis-1,3-Dichloropropene
 Concen: 10.046 ug/l
 RT: 10.087 min Scan# 1345
 Delta R.T. 0.012 min
 Lab File: VW031834.D
 Acq: 11 Jul 2025 20:56

Tgt Ion: 75 Resp: 142
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.4 38.2#
 39 38.8 41.8 62.6#





#55

1,1,2-Trichloroethane

Concen: 3.772 ug/l

RT: 10.770 min Scan# 14

Delta R.T. -0.018 min


Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :



Tgt Ion: 97 Resp: 28




Ion Ratio Lower Upper

97 100

83 0.0 66.9 100.3#

85 0.0 41.1 61.7#

99 0.0 46.6 69.8#



#56

Ethyl methacrylate

Concen: 2.582 ug/l

RT: 10.678 min Scan# 1442

Delta R.T. 0.031 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56




Tgt Ion: 69 Resp: 26


Ion Ratio Lower Upper

69 100

41 692.3 53.3 79.9#

39 188.5 30.2 45.2#





#57

1,3-Dichloropropane

Concen: 2.532 ug/l

RT: 10.934 min Scan# 1484

Delta R.T. 0.000 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 76 Resp: 33

Ion Ratio Lower Upper

76 100

78 0.0 26.2 39.4#



#58

2-Chloroethyl Vinyl ether

Concen: 3.426 ug/l

RT: 10.087 min Scan# 1345

Delta R.T. 0.159 min

Lab File: VW031834.D




Acq: 11 Jul 2025 20:56


Tgt Ion: 63 Resp: 19

Ion Ratio Lower Upper

63 100

106 0.0 24.1 36.1#





#59

2-Hexanone

Concen: 27.883 ug/l

RT: 10.983 min Scan# 1492

Delta R.T. 0.012 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 43 Resp: 145

Ion Ratio Lower Upper

43 100

58 0.0 27.9 83.6#



#61

1,2-Dibromoethane

Concen: 15.341 ug/l

RT: 11.245 min Scan# 1535

Delta R.T. 0.006 min

Lab File: VW031834.D




Acq: 11 Jul 2025 20:56


Tgt Ion: 107 Resp: 114

Ion Ratio Lower Upper

107 100

109 46.5 73.9 110.9#









#62
4-Bromofluorobenzene
Concen: 137.466 ug/l
RT: 12.617 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:	95	Resp:	1597
Ion	Ratio	Lower	Upper
95	100		
174	54.2	0.0	133.8
176	54.2	0.0	126.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.641 min Scan# 1600
Delta R.T. 0.012 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion:	117	Resp:	2012
Ion	Ratio	Lower	Upper
117	100		
82	69.2	48.6	72.8
119	34.2	23.9	35.9









#65
Chlorobenzene
Concen: 8.605 ug/l
RT: 11.666 min Scan# 11
Delta R.T. 0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:112 Resp: 382
Ion Ratio Lower Upper
112 100
114 0.0 23.8 35.8#



#67
Ethyl Benzene
Concen: 5.144 ug/l
RT: 11.733 min Scan# 1615
Delta R.T. 0.000 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 91 Resp: 396
Ion Ratio Lower Upper
91 100
106 0.0 25.4 38.2#









#68
m/p-Xylenes
Concen: 9.023 ug/l
RT: 11.855 min Scan# 1635
Delta R.T. 0.019 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:106 Resp: 267
Ion Ratio Lower Upper
106 100
91 361.8 166.4 249.6#



#69
o-Xylene
Concen: 3.832 ug/l
RT: 12.178 min Scan# 1688
Delta R.T. 0.012 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion:106 Resp: 105
Ion Ratio Lower Upper
106 100
91 546.7 111.5 334.4#





#70

Styrene

Concen: 11.443 ug/l

RT: 12.196 min Scan# 1

Delta R.T. 0.019 min


Lab File: VW031834.D

Acq: 11 Jul 2025 20:56




Instrument :

MSVOA_W

ClientSampleId :



Tgt Ion:	104	Resp:	542
Ion	Ratio	Lower	Upper
104	100		
78	40.2	42.1	63.1#
103	37.5	43.0	64.6#



#72

1,4-Dichlorobenzene-d4


Concen: 50.000 ug/l

RT: 13.556 min Scan# 1914



Delta R.T. 0.000 min


Lab File: VW031834.D

Acq: 11 Jul 2025 20:56



Tgt Ion:	152	Resp:	1520
Ion	Ratio	Lower	Upper
152	100		
115	83.4	31.9	95.7
150	172.1	0.0	356.4









#73
Isopropylbenzene
Concen: 2.067 ug/l
RT: 12.458 min Scan# 11
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:105 Resp: 239
Ion Ratio Lower Upper
105 100
120 0.0 12.8 38.4#



#74
N-ethyl acetate
Concen: 15.178 ug/l
RT: 12.294 min Scan# 1707
Delta R.T. 0.025 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 43 Resp: 473
Ion Ratio Lower Upper
43 100
70 10.6 36.0 54.0#
55 30.2 20.8 31.2
61 0.0 19.7 29.5#









#75
1,1,2,2-Tetrachloroethane
Concen: 7.402 ug/l
RT: 12.714 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 83 Resp: 190
Ion Ratio Lower Upper
83 100
131 0.0 5.3 16.1#
85 11.6 32.0 96.2#



#76
1,2,3-Trichloropropane
Concen: 3.055 ug/l
RT: 12.757 min Scan# 1783
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 75 Resp: 60
Ion Ratio Lower Upper
75 100
77 781.7 141.2 423.6#





#77

Bromobenzene

Concen: 4.847 ug/l

RT: 12.739 min Scan# 1781

Delta R.T. -0.006 min

Lab File: VW031834.D





Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :

Tgt Ion	Ratio	Lower	Upper
156	100		
77	375.2	117.9	353.7#
158	68.8	52.6	157.8



#78

n-propylbenzene

Concen: 4.110 ug/l




RT: 12.800 min Scan# 1790


Delta R.T. 0.000 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Tgt Ion	Ratio	Lower	Upper
91	100		
120	10.2	11.5	34.4#





#79

2-Chlorotoluene

Concen: 11.119 ug/l

RT: 13.001 min Scan# 1823

Delta R.T. 0.110 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 91 Resp: 927

Ion Ratio Lower Upper

91 100

126 16.6 15.8 47.4



#80

1,3,5-Trimethylbenzene

Concen: 4.817 ug/l

RT: 12.940 min Scan# 1813

Delta R.T. 0.000 min

Lab File: VW031834.D




Acq: 11 Jul 2025 20:56


Tgt Ion:105 Resp: 461

Ion Ratio Lower Upper

105 100

120 39.9 23.8 71.2









#81
trans-1,4-Dichloro-2-butene
Concen: 96.640 ug/l
RT: 12.623 min Scan# 1742
Delta R.T. 0.116 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 75 Resp: 815
Ion Ratio Lower Upper
75 100
53 0.0 83.5 125.3#
89 0.0 38.3 57.5#



#82
4-Chlorotoluene
Concen: 10.591 ug/l
RT: 13.001 min Scan# 1823
Delta R.T. 0.012 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56

Tgt Ion: 91 Resp: 927
Ion Ratio Lower Upper
91 100
126 16.6 16.0 47.9





#83

tert-Butylbenzene

Concen: 2.269 ug/l

RT: 13.196 min Scan# 1855

Delta R.T. -0.006 min

Lab File: VW031834.D





Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :

Tgt Ion:	119	Resp:	186
Ion	Ratio	Lower	Upper
119	100		
91	152.2	36.0	108.1#
134	0.0	12.4	37.2#



#84

1,2,4-Trimethylbenzene

Concen: 9.137 ug/l




RT: 13.245 min Scan# 1863


Delta R.T. 0.000 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Tgt Ion:	105	Resp:	886
Ion	Ratio	Lower	Upper
105	100		
120	47.1	22.1	66.1





#85

sec-Butylbenzene

Concen: 4.120 ug/l

RT: 13.385 min Scan# 11

Delta R.T. 0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion:105 Resp: 507

Ion Ratio Lower Upper

105 100

134 3.7 9.6 28.8#



#86

p-Isopropyltoluene

Concen: 4.871 ug/l

RT: 13.495 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56




Tgt Ion:119 Resp: 494


Ion Ratio Lower Upper

119 100

134 23.1 12.9 38.7

91 74.7 12.7 38.0#





#87

1,3-Dichlorobenzene

Concen: 7.870 ug/l

RT: 13.580 min Scan# 1918

Delta R.T. 0.086 min

Lab File: VW031834.D





Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :

Tgt Ion:146	Resp:	407
Ion Ratio	Lower	Upper
146	100	
111	50.1	22.3 66.9
148	84.3	31.8 95.4



#88

1,4-Dichlorobenzene

Concen: 7.689 ug/l




RT: 13.580 min Scan# 1918


Delta R.T. 0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Tgt Ion:146	Resp:	407
Ion Ratio	Lower	Upper
146	100	
111	50.1	20.4 61.3
148	84.3	30.2 90.6





#89

n-Butylbenzene

Concen: 5.156 ug/l

RT: 13.818 min Scan# 1957

Delta R.T. 0.000 min

Lab File: VW031834.D





Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :

Tgt Ion:	91	Resp:	508
Ion	Ratio	Lower	Upper
91	100		
92	54.7	27.2	81.5
134	4.3	12.2	36.6#



#90

Hexachloroethane

Concen: 2.514 ug/l




RT: 14.104 min Scan# 2004


Delta R.T. 0.019 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56


Tgt Ion:	117	Resp:	46
Ion	Ratio	Lower	Upper
117	100		
201	0.0	31.1	93.3#








#91
1,2-Dichlorobenzene
Concen: 8.145 ug/l
RT: 13.861 min Scan# 1964
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56


Instrument :
MSVOA_W
ClientSampleId :





Tgt Ion:146 Resp: 385
Ion Ratio Lower Upper
146 100
111 91.4 21.7 65.1#
148 58.2 29.8 89.5




#93
1,2,4-Trichlorobenzene
Concen: 10.189 ug/l
RT: 15.123 min Scan# 2171
Delta R.T. -0.006 min
Lab File: VW031834.D
Acq: 11 Jul 2025 20:56



Tgt Ion:180 Resp: 297
Ion Ratio Lower Upper
180 100
182 55.9 48.6 145.9
145 0.0 17.3 51.7#





#95

Naphthalene

Concen: 31.288 ug/l

RT: 15.366 min Scan# 2211

Delta R.T. 0.012 min

Lab File: VW031834.D





Acq: 11 Jul 2025 20:56

Instrument :

MSVOA_W

ClientSampleId :

Tgt Ion:128	Resp:	2236
Ion Ratio	Lower	Upper
128	100	
127	11.5	10.8 16.2
129	19.0	8.7 13.1#



#96

1,2,3-Trichlorobenzene

Concen: 15.531 ug/l

RT: 15.555 min Scan# 2242

Delta R.T. 0.006 min

Lab File: VW031834.D

Acq: 11 Jul 2025 20:56

Tgt Ion:180	Resp:	416
Ion Ratio	Lower	Upper
180	100	
182	67.1	46.4 139.1
145	0.0	17.8 53.3#

