




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071222\
Data File : VW023637.D
Acq On : 13 Jul 2022 05:12
Operator : SY/VA
Sample : N3678-15
Misc : 6.63g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F5B36

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/13/2022
Supervised By :Mahesh Dadoda 07/13/2022

Quant Time: Jul 13 06:16:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 13 02:42:59 2022
Response via : Initial Calibration



TIC: VW023637.D\data.ms

(25) Chloroform (T)

7.671min (+ 0.000) 0.69 ug/L

response 1392

Ion	Exp%	Act%
83.00	100.00	100.00
85.00	67.20	83.35
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071222\
Data File : VW023637.D
Acq On : 13 Jul 2022 05:12
Operator : SY/VA
Sample : N3678-15
Misc : 6.63g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F5B36

Manual IntegrationsAPPROVED

Quant Time: Jul 13 06:16:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 13 02:42:59 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 07/13/2022
Supervised By :Mahesh Dadoda 07/13/2022



TIC: VW023637.D\data.ms

(25) Chloroform (T)

7.671min (+ 0.000) 1.15 ug/L m

response 2326

Ion	Exp%	Act%
83.00	100.00	100.00
85.00	67.20	83.35
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071222\
 Data File : VW023637.D
 Acq On : 13 Jul 2022 05:12
 Operator : SY/VA
 Sample : N3678-15
 Misc : 6.63g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F5B36

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/13/2022
 Supervised By :Mahesh Dadoda 07/13/2022

Quant Time: Jul 13 06:16:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 13 02:42:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	70821	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	74139	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	26604	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	24195	18.873	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	75.480%	
7) Chloroethane-d5	2.879	69	20530	28.001	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	112.000%	
11) 1,1-Dichloroethene-d2	4.001	63	21831	15.104	ug/L	-0.02
Spiked Amount 25.000	Range 45	- 110	Recovery	=	60.400%	
21) 2-Butanone-d5	7.080	46	22067	82.890	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	165.780%#	
24) Chloroform-d	7.653	84	57184	28.348	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	113.400%	
26) 1,2-Dichloroethane-d4	8.305	65	39239	34.415	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	137.680%#	
32) Benzene-d6	8.275	84	97225	23.701	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	94.800%	
36) 1,2-Dichloropropane-d6	9.275	67	35957	29.940	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	119.760%	
41) Toluene-d8	10.323	98	79053	19.882	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	79.520%	
43) trans-1,3-Dichloroprop...	10.579	79	11121	20.744	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery	=	82.960%	
47) 2-Hexanone-d5	10.927	63	14450	61.518	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	123.040%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	37855	31.772	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	127.080%#	
66) 1,2-Dichlorobenzene-d4	13.853	152	26784	27.905	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	111.600%	
Target Compounds						
						Qvalue
13) Acetone	4.129	43	6986	35.606	ug/L	95
16) Methylene chloride	4.916	84	5380	4.479	ug/L	93
25) Chloroform	7.671	83	2326m	1.152	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071222\
Data File : VW023637.D
Acq On : 13 Jul 2022 05:12
Operator : SY/VA
Sample : N3678-15
Misc : 6.63g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F5B36

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/13/2022
Supervised By :Mahesh Dadoda 07/13/2022

Quant Time: Jul 13 06:16:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 13 02:42:59 2022
Response via : Initial Calibration

