

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071224\
 Data File : VW029555.D
 Acq On : 12 Jul 2024 16:44
 Operator : SY/MD
 Sample : P3230-11
 Misc : 5.24g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BHB53

Quant Time: Jul 12 23:27:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 12 23:17:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	507369	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	479384	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	230496	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	111143	13.560	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	54.240%		
7) Chloroethane-d5	2.893	69	86422	14.644	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	58.560%		
11) 1,1-Dichloroethene-d2	4.021	65	52518	14.269	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	57.080%		
21) 2-Butanone-d5	7.075	46	74834	30.416	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	60.840%		
24) Chloroform-d	7.648	84	233894	15.946	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	63.800%		
26) 1,2-Dichloroethane-d4	8.307	65	140022	16.275	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	65.120%#		
32) Benzene-d6	8.270	84	447437	15.580	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	62.320%		
36) 1,2-Dichloropropane-d6	9.276	67	149881	16.323	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	65.280%#		
41) Toluene-d8	10.319	98	398534	15.780	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	63.120%		
43) trans-1,3-Dichloroprop...	10.575	79	58523	15.051	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	60.200%		
47) 2-Hexanone-d5	10.922	63	57696	34.580	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	69.160%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	125648	16.854	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	67.400%		
66) 1,2-Dichlorobenzene-d4	13.855	152	133486	16.629	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	66.520%#		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	52190	5.746	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071224\
Data File : VW029555.D
Acq On : 12 Jul 2024 16:44
Operator : SY/MD
Sample : P3230-11
Misc : 5.24g/10mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BHB53

Quant Time: Jul 12 23:27:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 12 23:17:28 2024
Response via : Initial Calibration

