




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071322\
 Data File : VW023649.D
 Acq On : 13 Jul 2022 16:00
 Operator : SY/VA
 Sample : N3658-06
 Misc : 6.27g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AG7

Manual IntegrationsAPPROVED

Quant Time: Jul 14 02:13:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jul 14 02:04:07 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 07/14/2022
 Supervised By :Mahesh Dadoda 07/15/2022



TIC: VW023649.D\data.ms

(13) Acetone (T)

4.135min (+ 0.012) 2.74 ug/L

response 1754

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	32.10	38.77
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071322\
 Data File : VW023649.D
 Acq On : 13 Jul 2022 16:00
 Operator : SY/VA
 Sample : N3658-06
 Misc : 6.27g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AG7

Manual IntegrationsAPPROVED

Quant Time: Jul 14 02:13:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jul 14 02:04:07 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 07/14/2022
 Supervised By :Mahesh Dadoda 07/15/2022



TIC: VW023649.D\data.ms

(13) Acetone (T)

4.123min (-0.000) 5.97 ug/L m

response 3817

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	32.10	17.82
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071322\
Data File : VW023649.D
Acq On : 13 Jul 2022 16:00
Operator : SY/VA
Sample : N3658-06
Misc : 6.27g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AG7

Manual IntegrationsAPPROVED

Quant Time: Jul 14 02:13:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jul 14 02:04:07 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 07/14/2022
Supervised By :Mahesh Dadoda 07/15/2022



TIC: VW023649.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.573min (+ 0.268) 0.04 ug/L

response 165

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	49.70	52.12
51.00	92.20	88.48
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071322\
 Data File : VW023649.D
 Acq On : 13 Jul 2022 16:00
 Operator : SY/VA
 Sample : N3658-06
 Misc : 6.27g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AG7

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/14/2022
 Supervised By :Mahesh Dadoda 07/15/2022

Quant Time: Jul 14 02:13:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jul 14 02:04:07 2022
 Response via : Initial Calibration



TIC: VW023649.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.305min (-0.000) 24.12 ug/L m

response 98016

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	49.70	0.09#
51.00	92.20	0.15#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071322\
 Data File : VW023649.D
 Acq On : 13 Jul 2022 16:00
 Operator : SY/VA
 Sample : N3658-06
 Misc : 6.27g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AG7

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/14/2022
 Supervised By :Mahesh Dadoda 07/15/2022

Quant Time: Jul 14 02:13:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jul 14 02:04:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	228223	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	225283	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	107022	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	145022	24.879	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	99.520%	
7) Chloroethane-d5	2.879	69	99233	25.519	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	102.080%	
11) 1,1-Dichloroethene-d2	4.013	63	79022	16.648	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery	=	66.600%	
21) 2-Butanone-d5	7.080	46	50824	56.213	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	112.420%	
24) Chloroform-d	7.647	84	156737	23.588	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	94.360%	
26) 1,2-Dichloroethane-d4	8.305	65	98016m	24.119	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	96.480%	
32) Benzene-d6	8.275	84	290731	24.826	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	99.320%	
36) 1,2-Dichloropropane-d6	9.274	67	90071	25.534	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	102.120%	
41) Toluene-d8	10.323	98	279725	23.549	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	94.200%	
43) trans-1,3-Dichloroprop...	10.573	79	38222	23.116	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery	=	92.480%	
47) 2-Hexanone-d5	10.921	63	42798	57.697	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	115.400%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	93679	24.030	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	96.120%	
66) 1,2-Dichlorobenzene-d4	13.853	152	96609	25.287	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	101.160%	
Target Compounds						
13) Acetone	4.123	43	3817m	5.971	ug/L	
20) cis-1,2-Dichloroethene	7.165	96	52928	16.276	ug/L	100
34) Trichloroethene	9.092	95	36948	10.929	ug/L	96
46) Tetrachloroethene	10.860	164	407114	155.742	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071322\
Data File : VW023649.D
Acq On : 13 Jul 2022 16:00
Operator : SY/VA
Sample : N3658-06
Misc : 6.27g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AG7

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/14/2022
Supervised By :Mahesh Dadoda 07/15/2022

Quant Time: Jul 14 02:13:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jul 14 02:04:07 2022
Response via : Initial Calibration

