

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071323\
 Data File : VW026607.D
 Acq On : 13 Jul 2023 09:34
 Operator : SY/MD
 Sample : VW0713SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK539

Quant Time: Jul 14 06:26:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 14 06:26:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	671806	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	619358	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	268958	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	227525	20.078	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.320%
7) Chloroethane-d5	2.893	69	163648	20.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.320%
11) 1,1-Dichloroethene-d2	4.021	65	106831	20.074	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	80.280%
21) 2-Butanone-d5	7.075	46	159954	43.675	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.360%
24) Chloroform-d	7.648	84	434568	22.468	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.880%
26) 1,2-Dichloroethane-d4	8.307	65	254800	23.594	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.360%
32) Benzene-d6	8.276	84	904735	21.279	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.120%
36) 1,2-Dichloropropane-d6	9.270	67	286254	21.896	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.600%
41) Toluene-d8	10.325	98	769546	21.103	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.400%
43) trans-1,3-Dichloroprop...	10.581	79	110339	21.061	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.240%
47) 2-Hexanone-d5	10.922	63	112532	44.113	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.220%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	241001	21.769	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.080%
66) 1,2-Dichlorobenzene-d4	13.848	152	247472	25.061	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.240%
Target Compounds						
16) Methylene chloride	4.923	84	25246	2.127	ug/L	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071323\
Data File : VW026607.D
Acq On : 13 Jul 2023 09:34
Operator : SY/MD
Sample : VW0713SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK539

Quant Time: Jul 14 06:26:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 14 06:26:25 2023
Response via : Initial Calibration

