

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071422\
 Data File : VW023727.D
 Acq On : 15 Jul 2022 05:14
 Operator : SY/VA
 Sample : N3700-13
 Misc : 6.24g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F5C20

Quant Time: Jul 15 06:30:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 15 06:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	167127	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	160880	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	59933	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	61986	14.522	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.080%
7) Chloroethane-d5	2.879	69	52094	18.294	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.160%
11) 1,1-Dichloroethene-d2	4.013	63	38853	11.178	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	44.720%#
21) 2-Butanone-d5	7.074	46	35541	53.679	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.360%
24) Chloroform-d	7.647	84	109644	22.533	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.120%
26) 1,2-Dichloroethane-d4	8.305	65	71579	24.052	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	8.275	84	187288	22.395	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.560%
36) 1,2-Dichloropropane-d6	9.274	67	64765	25.709	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.840%
41) Toluene-d8	10.323	98	166532	19.632	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.520%
43) trans-1,3-Dichloroprop...	10.573	79	20572	17.422	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.680%
47) 2-Hexanone-d5	10.921	63	25590	48.309	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.620%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	60824	21.848	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.400%
66) 1,2-Dichlorobenzene-d4	13.853	152	48116	22.490	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.960%
Target Compounds						
13) Acetone	4.129	43	14020	29.950	ug/L	Qvalue 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071422\
Data File : VW023727.D
Acq On : 15 Jul 2022 05:14
Operator : SY/VA
Sample : N3700-13
Misc : 6.24g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F5C20

Quant Time: Jul 15 06:30:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 15 06:03:21 2022
Response via : Initial Calibration

