

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071422\
 Data File : VW023723.D
 Acq On : 15 Jul 2022 01:50
 Operator : SY/VA
 Sample : VW0714SBL03
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK547

Quant Time: Jul 15 09:57:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 15 06:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	164830	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	166708	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	74389	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	126338	30.010	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	120.040%
7) Chloroethane-d5	2.879	69	67986	24.207	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.840%
11) 1,1-Dichloroethene-d2	4.007	63	56170	16.385	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.560%
21) 2-Butanone-d5	7.074	46	28760	44.043	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.080%
24) Chloroform-d	7.647	84	109540	22.826	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.320%
26) 1,2-Dichloroethane-d4	8.305	65	68329	23.280	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.120%
32) Benzene-d6	8.275	84	206915	23.877	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.520%
36) 1,2-Dichloropropane-d6	9.274	67	62716	24.026	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.120%
41) Toluene-d8	10.323	98	199628	22.710	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.840%
43) trans-1,3-Dichloroprop...	10.579	79	24723	20.205	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.840%
47) 2-Hexanone-d5	10.927	63	23249	42.355	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.720%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	58183	20.169	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	64652	24.346	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071422\
Data File : VW023723.D
Acq On : 15 Jul 2022 01:50
Operator : SY/VA
Sample : VW0714SBL03
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK547

Quant Time: Jul 15 09:57:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 15 06:03:21 2022
Response via : Initial Calibration

