

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071425\
 Data File : VW031852.D
 Acq On : 14 Jul 2025 16:14
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK

Quant Time: Jul 15 02:46:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	188491	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.856	114	405265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.636	117	374818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	186601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.325	65	135211	49.985	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	99.960%
35) Dibromofluoromethane	7.904	113	120882	45.710	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	91.420%
50) Toluene-d8	10.331	98	442496	44.963	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	89.920%
62) 4-Bromofluorobenzene	12.617	95	182426	50.371	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	100.740%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.960	84	27792	6.491	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071425\
Data File : VW031852.D
Acq On : 14 Jul 2025 16:14
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK

Quant Time: Jul 15 02:46:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:35:17 2025
Response via : Initial Calibration

