

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071425\
 Data File : VW031853.D
 Acq On : 14 Jul 2025 16:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025
 Supervised By :Mahesh Dadoda 07/16/2025

Quant Time: Jul 15 02:46:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	212729	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	424856	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	377830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	182654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	168367	55.150	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery = 110.300%			
35) Dibromofluoromethane	7.910	113	136117	49.098	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery = 98.200%			
50) Toluene-d8	10.324	98	505955	49.040	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery = 98.080%			
62) 4-Bromofluorobenzene	12.617	95	196126	51.657	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery = 103.320%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.051	85	57333	39.574	ug/l	94
3) Chloromethane	2.265	50	86152	49.325	ug/l	99
4) Vinyl Chloride	2.411	62	109868	47.858	ug/l	97
5) Bromomethane	2.832	94	78374	43.707	ug/l	96
6) Chloroethane	2.978	64	72431	46.447	ug/l	99
7) Trichlorofluoromethane	3.313	101	102939	49.373	ug/l	99
8) Diethyl Ether	3.728	74	85658	54.068	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.106	101	108819	47.023	ug/l	99
10) Methyl Iodide	4.319	142	165157	46.294	ug/l	98
11) Tert butyl alcohol	5.228	59	51725	273.480	ug/l	97
12) 1,1-Dichloroethene	4.082	96	121853	48.072	ug/l	96
13) Acrolein	3.935	56	67079	197.326	ug/l	99
14) Allyl chloride	4.709	41	196006	50.398	ug/l	96
15) Acrylonitrile	5.411	53	211168	271.049	ug/l	99
16) Acetone	4.167	43	158259	209.730	ug/l	100
17) Carbon Disulfide	4.429	76	292564	43.132	ug/l	97
18) Methyl Acetate	4.716	43	125948	58.402	ug/l	99
19) Methyl tert-butyl Ether	5.471	73	246910	57.390	ug/l	99
20) Methylene Chloride	4.965	84	159316	53.351	ug/l	97
21) trans-1,2-Dichloroethene	5.465	96	135951	50.476	ug/l	92
22) Diisopropyl ether	6.343	45	446479	58.863	ug/l	98
23) Vinyl Acetate	6.288	43	1483523	286.345	ug/l	98
24) 1,1-Dichloroethane	6.252	63	264279	53.742	ug/l	99
25) 2-Butanone	7.197	43	278466	283.659	ug/l	97
26) 2,2-Dichloropropane	7.191	77	129101	44.996	ug/l	98
27) cis-1,2-Dichloroethene	7.197	96	170957	55.218	ug/l	99
28) Bromochloromethane	7.532	49	120663	55.603	ug/l	97
29) Tetrahydrofuran	7.556	42	193597	287.967	ug/l	99
30) Chloroform	7.697	83	287058	54.936	ug/l	96
31) Cyclohexane	7.971	56	203663	46.852	ug/l	95
32) 1,1,1-Trichloroethane	7.892	97	199562	49.558	ug/l	99
36) 1,1-Dichloropropene	8.099	75	188890	47.084	ug/l	100
37) Ethyl Acetate	7.276	43	131578	51.993	ug/l	97
38) Carbon Tetrachloride	8.081	117	176848	43.259	ug/l	97
39) Methylcyclohexane	9.343	83	236145	47.311	ug/l	99
40) Benzene	8.337	78	590167	49.723	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071425\
 Data File : VW031853.D
 Acq On : 14 Jul 2025 16:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 07/15/2025
 Supervised By : Mahesh Dadoda 07/16/2025

Quant Time: Jul 15 02:46:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.502	41	72747	53.661	ug/l	98
42) 1,2-Dichloroethane	8.416	62	200375	49.963	ug/l	99
43) Isopropyl Acetate	8.440	43	231733	53.121	ug/l	99
44) Trichloroethene	9.105	130	138308	46.687	ug/l	98
45) 1,2-Dichloropropane	9.379	63	145411	50.690	ug/l	96
46) Dibromomethane	9.465	93	93543	50.509	ug/l	96
47) Bromodichloromethane	9.654	83	219148	50.461	ug/l	96
48) Methyl methacrylate	9.446	41	114937	53.583	ug/l	97
49) 1,4-Dioxane	9.465	88	22880	1054.828	ug/l #	97
51) 4-Methyl-2-Pentanone	10.215	43	666519	265.760	ug/l	99
52) Toluene	10.391	92	370626	48.770	ug/l	100
53) t-1,3-Dichloropropene	10.611	75	212988	52.270	ug/l	98
54) cis-1,3-Dichloropropene	10.080	75	238250	51.577	ug/l	97
55) 1,1,2-Trichloroethane	10.788	97	124970	51.513	ug/l	95
56) Ethyl methacrylate	10.647	69	190816	57.977	ug/l	98
57) 1,3-Dichloropropane	10.934	76	224026	52.597	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.928	63	488946	269.792	ug/l	97
59) 2-Hexanone	10.970	43	465349	273.814	ug/l	99
60) Dibromochloromethane	11.129	129	144104	49.673	ug/l	97
61) 1,2-Dibromoethane	11.239	107	124688	51.342	ug/l	100
64) Tetrachloroethene	10.867	164	109175	44.174	ug/l #	85
65) Chlorobenzene	11.659	112	418496	50.202	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.726	131	131225	49.398	ug/l	98
67) Ethyl Benzene	11.733	91	732870	50.699	ug/l	98
68) m/p-Xylenes	11.842	106	550032	98.983	ug/l	100
69) o-Xylene	12.165	106	261949	50.913	ug/l	97
70) Styrene	12.178	104	465431	52.327	ug/l	100
71) Bromoform	12.348	173	77558	48.940	ug/l #	99
73) Isopropylbenzene	12.464	105	673099	48.446	ug/l	100
74) N-amyl acetate	12.269	43	213154	56.920	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.714	83	157252	50.981	ug/l	99
76) 1,2,3-Trichloropropane	12.763	75	121986m	51.680	ug/l	
77) Bromobenzene	12.745	156	151496	48.889	ug/l	94
78) n-propylbenzene	12.799	91	813177	48.877	ug/l	99
79) 2-Chlorotoluene	12.891	91	506996	50.606	ug/l	98
80) 1,3,5-Trimethylbenzene	12.940	105	591840	51.462	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.513	75	49593	48.937	ug/l	97
82) 4-Chlorotoluene	12.988	91	529953	50.384	ug/l	99
83) tert-Butylbenzene	13.202	119	478225	48.545	ug/l	100
84) 1,2,4-Trimethylbenzene	13.251	105	570575	48.966	ug/l	100
85) sec-Butylbenzene	13.379	105	735351	49.732	ug/l	98
86) p-Isopropyltoluene	13.494	119	607316	49.837	ug/l	98
87) 1,3-Dichlorobenzene	13.494	146	302447	48.669	ug/l	99
88) 1,4-Dichlorobenzene	13.580	146	289785	45.559	ug/l	93
89) n-Butylbenzene	13.818	91	581390	49.104	ug/l	99
90) Hexachloroethane	14.086	117	103878	47.250	ug/l	96
91) 1,2-Dichlorobenzene	13.866	146	275295	48.468	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.482	75	28754	47.742	ug/l	95
93) 1,2,4-Trichlorobenzene	15.128	180	174369	49.778	ug/l	94
94) Hexachlorobutadiene	15.232	225	74794	44.218	ug/l	90
95) Naphthalene	15.360	128	460711	53.647	ug/l	99
96) 1,2,3-Trichlorobenzene	15.549	180	159265	49.482	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071425\
Data File : VW031853.D
Acq On : 14 Jul 2025 16:37
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025

Quant Time: Jul 15 02:46:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:35:17 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071425\
 Data File : VW031853.D
 Acq On : 14 Jul 2025 16:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1


Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jul 15 02:46:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
 Supervised By :Mahesh Dadoda 07/16/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.965 min Scan# 99
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 212729
Ion Ratio Lower Upper
168 100
99 74.6 49.4 74.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#2
Dichlorodifluoromethane
Concen: 39.574 ug/l
RT: 2.051 min Scan# 27
Delta R.T. 0.012 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion: 85 Resp: 57333
Ion Ratio Lower Upper
85 100
87 34.1 15.4 46.2









#3
Chloromethane
Concen: 49.325 ug/l
RT: 2.265 min Scan# 61
Delta R.T. 0.012 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 50 Resp: 86152
Ion Ratio Lower Upper
50 100
52 32.0 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#4
Vinyl Chloride
Concen: 47.858 ug/l
RT: 2.411 min Scan# 86
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 62 Resp: 109868
Ion Ratio Lower Upper
62 100
64 31.7 26.8 40.2





#5
Bromomethane
Concen: 43.707 ug/l
RT: 2.832 min Scan# 11
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 78374
Ion Ratio Lower Upper
94 100
96 97.9 75.6 113.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#6
Chloroethane
Concen: 46.447 ug/l
RT: 2.978 min Scan# 179
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion: 64 Resp: 72431
Ion Ratio Lower Upper
64 100
66 31.0 24.5 36.7





#7

Trichlorofluoromethane

Concen: 49.373 ug/l

RT: 3.313 min Scan# 21

Delta R.T. 0.006 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC

Tgt Ion:101 Resp: 102939

Ion Ratio Lower Upper

101 100





103 63.9 50.2 75.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025

Supervised By :Mahesh Dadoda 07/16/2025



#8

Diethyl Ether

Concen: 54.068 ug/l

RT: 3.728 min Scan# 302

Delta R.T. 0.006 min

Lab File: VW031853.D




Acq: 14 Jul 2025 16:37


Tgt Ion: 74 Resp: 85658

Ion Ratio Lower Upper

74 100


45 91.8 43.9 131.7





#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.023 ug/l
RT: 4.106 min Scan# 301
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion:101 Resp: 108819
Ion Ratio Lower Upper
101 100
85 44.7 35.6 53.4
151 69.3 56.8 85.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#10
Methyl Iodide
Concen: 46.294 ug/l
RT: 4.319 min Scan# 399
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37



Tgt Ion:142 Resp: 165157
Ion Ratio Lower Upper
142 100
127 46.2 37.9 56.9
141 14.3 11.7 17.5









#11
Tert butyl alcohol
Concen: 273.480 ug/l
RT: 5.228 min Scan# 546
Delta R.T. 0.012 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 59 Resp: 51729
Ion Ratio Lower Upper
59 100
57 10.6 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#12
1,1-Dichloroethene
Concen: 48.072 ug/l
RT: 4.082 min Scan# 360
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion: 96 Resp: 121853
Ion Ratio Lower Upper
96 100
61 162.2 135.8 203.8
98 63.2 50.4 75.6





#13

Acrolein

Concen: 197.326 ug/l

RT: 3.935 min Scan# 335

Delta R.T. 0.006 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 67079

Ion Ratio Lower Upper

56 100





55 68.2 54.0 81.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025

Supervised By :Mahesh Dadoda 07/16/2025



#14

Allyl chloride

Concen: 50.398 ug/l

RT: 4.709 min Scan# 463

Delta R.T. 0.006 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37




Tgt Ion: 41 Resp: 196006


Ion Ratio Lower Upper

41 100

39 71.4 60.7 91.1


76 37.3 30.3 45.5





#15
Acrylonitrile
Concen: 271.049 ug/l
RT: 5.411 min Scan# 511
Delta R.T. 0.012 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 211168
Ion Ratio Lower Upper
53 100
52 83.2 66.4 99.6
51 36.6 29.9 44.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#16
Acetone
Concen: 209.730 ug/l
RT: 4.167 min Scan# 374
Delta R.T. 0.012 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion: 43 Resp: 158259
Ion Ratio Lower Upper
43 100
58 30.8 24.8 37.2









#17
Carbon Disulfide
Concen: 43.132 ug/l
RT: 4.429 min Scan# 416
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 76 Resp: 292564
Ion Ratio Lower Upper
76 100
78 9.3 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#18
Methyl Acetate
Concen: 58.402 ug/l
RT: 4.716 min Scan# 464
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 43 Resp: 125948
Ion Ratio Lower Upper
43 100
74 25.1 20.6 30.8





#19
Methyl tert-butyl Ether
Concen: 57.390 ug/l
RT: 5.471 min Scan# 505
Delta R.T. 0.012 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 73 Resp: 246910
Ion Ratio Lower Upper
73 100
57 21.9 17.2 25.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#20
Methylene Chloride
Concen: 53.351 ug/l
RT: 4.965 min Scan# 505
Delta R.T. 0.012 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion: 84 Resp: 159316
Ion Ratio Lower Upper
84 100
49 119.1 98.2 147.4
51 36.5 29.8 44.8
86 64.4 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 50.476 ug/l
RT: 5.465 min Scan# 586
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 13595
Ion Ratio Lower Upper
96 100
61 127.6 109.8 164.6
98 61.1 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#22
Diisopropyl ether
Concen: 58.863 ug/l
RT: 6.343 min Scan# 731
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 45 Resp: 446479
Ion Ratio Lower Upper
45 100
43 61.2 48.3 72.5
87 26.2 22.6 34.0
59 11.8 9.7 14.5





#23
 Vinyl Acetate
 Concen: 286.345 ug/l
 RT: 6.288 min Scan# 716
 Delta R.T. 0.006 min
 Lab File: VW031853.D
 Acq: 14 Jul 2025 16:37




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 43 Resp: 148352
 Ion Ratio Lower Upper
 43 100
 86 10.6 9.2 13.8

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
 Supervised By :Mahesh Dadoda 07/16/2025




#24
 1,1-Dichloroethane
 Concen: 53.742 ug/l
 RT: 6.252 min Scan# 716
 Delta R.T. 0.006 min
 Lab File: VW031853.D
 Acq: 14 Jul 2025 16:37




Tgt Ion: 63 Resp: 264279
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.5 10.6
 100 4.4 2.5 7.5





#25
2-Butanone
Concen: 283.659 ug/l
RT: 7.197 min Scan# 81
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 27846
Ion Ratio Lower Upper
43 100
72 27.3 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#26
2,2-Dichloropropane
Concen: 44.996 ug/l
RT: 7.191 min Scan# 870
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion: 77 Resp: 129101
Ion Ratio Lower Upper
77 100
97 24.3 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 55.218 ug/l

RT: 7.197 min Scan# 871

Delta R.T. 0.006 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 96	Resp: 17095
Ion Ratio	Lower Upper
96 100	
61 142.0	0.0 286.8
98 65.1	0.0 130.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#28

Bromochloromethane

Concen: 55.603 ug/l




RT: 7.532 min Scan# 926


Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37


Tgt Ion: 49	Resp: 120663
Ion Ratio	Lower Upper
49 100	
129 1.7	0.0 2.4
130 73.9	61.3 91.9





#29
Tetrahydrofuran
Concen: 287.967 ug/l
RT: 7.556 min Scan# 91
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 42 Resp: 19359
Ion Ratio Lower Upper
42 100
72 45.7 36.8 55.2
71 42.9 34.6 52.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#30
Chloroform
Concen: 54.936 ug/l
RT: 7.697 min Scan# 953
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 83 Resp: 287058
Ion Ratio Lower Upper
83 100
85 67.2 51.4 77.0





#31
Cyclohexane
Concen: 46.852 ug/l
RT: 7.971 min Scan# 998
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 20366
Ion Ratio Lower Upper
56 100
69 36.4 27.2 40.8
84 89.9 68.5 102.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#32
1,1,1-Trichloroethane
Concen: 49.558 ug/l
RT: 7.892 min Scan# 985
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion: 97 Resp: 199562
Ion Ratio Lower Upper
97 100
99 63.5 51.0 76.6
61 44.3 36.0 54.0





#33
1,2-Dichloroethane-d4
Concen: 55.150 ug/l
RT: 8.319 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 65 Resp: 16836
Ion Ratio Lower Upper
65 100
67 51.1 0.0 99.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.855 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion:114 Resp: 424856
Ion Ratio Lower Upper
114 100
63 22.0 0.0 43.6
88 16.0 0.0 34.2





#35
 Dibromofluoromethane
 Concen: 49.098 ug/l
 RT: 7.910 min Scan# 987
 Delta R.T. 0.006 min
 Lab File: VW031853.D
 Acq: 14 Jul 2025 16:37




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 113 Resp: 13611
 Ion Ratio Lower Upper
 113 100
 111 103.8 82.1 123.1
 192 16.2 13.8 20.6




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
 Supervised By :Mahesh Dadoda 07/16/2025



#36
 1,1-Dichloropropene
 Concen: 47.084 ug/l
 RT: 8.099 min Scan# 1019
 Delta R.T. 0.006 min
 Lab File: VW031853.D
 Acq: 14 Jul 2025 16:37


Tgt Ion: 75 Resp: 188890
 Ion Ratio Lower Upper
 75 100
 110 33.5 16.8 50.4
 77 31.5 25.0 37.6





#37
Ethyl Acetate
Concen: 51.993 ug/l
RT: 7.276 min Scan# 884
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 131578
Ion Ratio Lower Upper
43 100
61 12.4 10.9 16.3
70 10.6 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#38
Carbon Tetrachloride
Concen: 43.259 ug/l
RT: 8.081 min Scan# 1016
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion:117 Resp: 176848
Ion Ratio Lower Upper
117 100
119 96.0 75.0 112.6
121 27.8 24.6 37.0






#39
Methylcyclohexane
Concen: 47.311 ug/l
RT: 9.343 min Scan# 1223
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 83 Resp: 236149
Ion Ratio Lower Upper
83 100
55 76.5 61.0 91.4
98 46.8 37.9 56.9

Manual Integrations




APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#40
Benzene
Concen: 49.723 ug/l
RT: 8.337 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 78 Resp: 590167
Ion Ratio Lower Upper
78 100
77 23.9 19.2 28.8





#41
Methacrylonitrile
Concen: 53.661 ug/l
RT: 7.502 min Scan# 91
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 7274
Ion Ratio Lower Upper
41 100
39 55.2 43.2 64.8
67 75.0 58.1 87.1
52 32.2 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#42
1,2-Dichloroethane
Concen: 49.963 ug/l
RT: 8.416 min Scan# 1071
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 62 Resp: 200375
Ion Ratio Lower Upper
62 100
98 9.1 0.0 17.2





#43
Isopropyl Acetate
Concen: 53.121 ug/l
RT: 8.440 min Scan# 1075
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 231733
Ion Ratio Lower Upper
43 100
61 29.6 23.1 34.7
87 13.6 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#44
Trichloroethene
Concen: 46.687 ug/l
RT: 9.105 min Scan# 1184
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion:130 Resp: 138308
Ion Ratio Lower Upper
130 100
95 106.8 0.0 208.6





#45
1,2-Dichloropropane
Concen: 50.690 ug/l
RT: 9.379 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 63 Resp: 14541
Ion Ratio Lower Upper
63 100
65 32.5 24.2 36.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#46
Dibromomethane
Concen: 50.509 ug/l
RT: 9.465 min Scan# 1243
Delta R.T. -0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 93 Resp: 93543
Ion Ratio Lower Upper
93 100
95 83.3 67.0 100.4
174 83.4 71.7 107.5





#47
Bromodichloromethane
Concen: 50.461 ug/l
RT: 9.654 min Scan# 1274
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 21914
Ion Ratio Lower Upper
83 100
85 66.4 50.8 76.2
127 7.6 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#48
Methyl methacrylate
Concen: 53.583 ug/l
RT: 9.446 min Scan# 1240
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion: 41 Resp: 114937
Ion Ratio Lower Upper
41 100
69 88.4 71.6 107.4
39 57.6 49.5 74.3





#49

1,4-Dioxane

Concen: 1054.828 ug/l

RT: 9.465 min Scan# 1243

Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 22880

Ion Ratio Lower Upper

88 100

43 23.6 0.0 0.0





58 75.5 58.3 87.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025

Supervised By :Mahesh Dadoda 07/16/2025



#50

Toluene-d8

Concen: 49.040 ug/l

RT: 10.324 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VW031853.D




Acq: 14 Jul 2025 16:37


Tgt Ion: 98 Resp: 505955

Ion Ratio Lower Upper

98 100


100 65.5 53.0 79.4





#51
4-Methyl-2-Pentanone
Concen: 265.760 ug/l
RT: 10.215 min Scan# 1366
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 666519
Ion Ratio Lower Upper
43 100
58 40.6 32.2 48.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#52
Toluene
Concen: 48.770 ug/l
RT: 10.391 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 92 Resp: 370626
Ion Ratio Lower Upper
92 100
91 174.6 139.5 209.3





#53
t-1,3-Dichloropropene
Concen: 52.270 ug/l
RT: 10.611 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 212988
Ion Ratio Lower Upper
75 100
77 31.6 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#54
cis-1,3-Dichloropropene
Concen: 51.577 ug/l
RT: 10.080 min Scan# 1344
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 75 Resp: 238250
Ion Ratio Lower Upper
75 100
77 32.8 25.4 38.2
39 49.8 41.8 62.6





#55
1,1,2-Trichloroethane
Concen: 51.513 ug/l
RT: 10.788 min Scan# 1460
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion	Ratio	Lower	Upper
97	100		
83	87.9	66.9	100.3
85	54.9	41.1	61.7
99	61.7	46.6	69.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#56
Ethyl methacrylate
Concen: 57.977 ug/l
RT: 10.647 min Scan# 1437
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion	Ratio	Lower	Upper
69	100		
41	65.8	53.3	79.9
39	35.1	30.2	45.2





#57
1,3-Dichloropropane
Concen: 52.597 ug/l
RT: 10.934 min Scan# 1484
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 224020
Ion Ratio Lower Upper
76 100
78 31.8 26.2 39.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#58
2-Chloroethyl Vinyl ether
Concen: 269.792 ug/l
RT: 9.928 min Scan# 1319
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 63 Resp: 488946
Ion Ratio Lower Upper
63 100
106 28.6 24.1 36.1





#59
2-Hexanone
Concen: 273.814 ug/l
RT: 10.970 min Scan# 1490
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 465349
Ion Ratio Lower Upper
43 100
58 56.7 27.9 83.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#60
Dibromochloromethane
Concen: 49.673 ug/l
RT: 11.129 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion:129 Resp: 144104
Ion Ratio Lower Upper
129 100
127 79.0 38.0 114.1





#61
1,2-Dibromoethane
Concen: 51.342 ug/l
RT: 11.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 124688
Ion Ratio Lower Upper
107 100
109 92.5 73.9 110.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#62
4-Bromofluorobenzene
Concen: 51.657 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37



Tgt Ion: 95 Resp: 196126
Ion Ratio Lower Upper
95 100
174 63.5 0.0 133.8
176 63.1 0.0 126.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.629 min Scan# 11

Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC

Tgt Ion:117 Resp: 377830





Ion	Ratio	Lower	Upper
117	100		
82	59.4	48.6	72.8
119	31.8	23.9	35.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025

Supervised By :Mahesh Dadoda 07/16/2025



#64

Tetrachloroethene

Concen: 44.174 ug/l

RT: 10.867 min Scan# 1473




Delta R.T. -0.000 min


Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Tgt Ion:164 Resp: 109175


Ion	Ratio	Lower	Upper
164	100		
166	137.1	100.7	151.1
129	114.1	75.4	113.0#
131	109.4	74.0	111.0





#65
Chlorobenzene
Concen: 50.202 ug/l
RT: 11.659 min Scan# 1160
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 418490
Ion Ratio Lower Upper
112 100
114 30.4 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 49.398 ug/l
RT: 11.726 min Scan# 1614
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion:131 Resp: 131225
Ion Ratio Lower Upper
131 100
133 94.4 48.5 145.6
119 65.7 32.5 97.5





#67
Ethyl Benzene
Concen: 50.699 ug/l
RT: 11.733 min Scan# 1615
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 732870
Ion Ratio Lower Upper
91 100
106 30.7 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#68
m/p-Xylenes
Concen: 98.983 ug/l
RT: 11.842 min Scan# 1633
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion:106 Resp: 550032
Ion Ratio Lower Upper
106 100
91 208.4 166.4 249.6





#69
o-Xylene
Concen: 50.913 ug/l
RT: 12.165 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion:106 Resp: 261949
Ion Ratio Lower Upper
106 100
91 227.2 111.5 334.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#70
Styrene
Concen: 52.327 ug/l
RT: 12.178 min Scan# 1688
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Tgt Ion:104 Resp: 465431
Ion Ratio Lower Upper
104 100
78 53.1 42.1 63.1
103 53.7 43.0 64.6





#71

Bromoform

Concen: 48.940 ug/l

RT: 12.348 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VW031853.D


Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC






Tgt Ion:	173	Resp:	7755
Ion Ratio	Lower	Upper	
173	100		
175	48.8	24.0	72.0
254	0.1	0.1	0.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.555 min Scan# 1914


Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Tgt Ion:	152	Resp:	182654
Ion Ratio	Lower	Upper	
152	100		
115	63.7	31.9	95.7
150	171.7	0.0	356.4









#73
Isopropylbenzene
Concen: 48.446 ug/l
RT: 12.464 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37

Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC

Tgt Ion:105 Resp: 673099
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#74
N-ethyl acetate
Concen: 56.920 ug/l
RT: 12.269 min Scan# 1703
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 43 Resp: 213154
Ion Ratio Lower Upper
43 100
70 45.3 36.0 54.0
55 24.8 20.8 31.2
61 24.7 19.7 29.5





#75
1,1,2,2-Tetrachloroethane
Concen: 50.981 ug/l
RT: 12.714 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 83 Resp: 157252
Ion Ratio Lower Upper
83 100
131 10.2 5.3 16.1
85 64.6 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#76
1,2,3-Trichloropropane
Concen: 51.680 ug/l m
RT: 12.763 min Scan# 1784
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37



Tgt Ion: 75 Resp: 121986
Ion Ratio Lower Upper
75 100
77 282.8 141.2 423.6





#77

Bromobenzene

Concen: 48.889 ug/l

RT: 12.745 min Scan# 1781

Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :





VSTDCCC050EC

Tgt Ion:156 Resp: 151490
 Ion Ratio Lower Upper
 156 100
 77 227.7 117.9 353.7
 158 96.7 52.6 157.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025
 Supervised By :Mahesh Dadoda 07/16/2025



#78

n-propylbenzene

Concen: 48.877 ug/l




RT: 12.799 min Scan# 1790


Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37


Tgt Ion: 91 Resp: 813177
 Ion Ratio Lower Upper
 91 100
 120 22.2 11.5 34.4





#79
2-Chlorotoluene
Concen: 50.606 ug/l
RT: 12.891 min Scan# 1805
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 506990
Ion Ratio Lower Upper
91 100
126 30.5 15.8 47.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#80
1,3,5-Trimethylbenzene
Concen: 51.462 ug/l
RT: 12.940 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion:105 Resp: 591840
Ion Ratio Lower Upper
105 100
120 46.2 23.8 71.2





#81
trans-1,4-Dichloro-2-butene
Concen: 48.937 ug/l
RT: 12.513 min Scan# 1743
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 4959
Ion Ratio Lower Upper
75 100
53 108.8 83.5 125.3
89 47.0 38.3 57.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#82
4-Chlorotoluene
Concen: 50.384 ug/l
RT: 12.988 min Scan# 1821
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37


Tgt Ion: 91 Resp: 529953
Ion Ratio Lower Upper
91 100
126 31.2 16.0 47.9





#83
tert-Butylbenzene
Concen: 48.545 ug/l
RT: 13.202 min Scan# 1856
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 478225
Ion Ratio Lower Upper
119 100
91 72.5 36.0 108.1
134 24.7 12.4 37.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 48.966 ug/l
RT: 13.251 min Scan# 1864
Delta R.T. 0.006 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion:105 Resp: 570575
Ion Ratio Lower Upper
105 100
120 43.8 22.1 66.1





#85
 sec-Butylbenzene
 Concen: 49.732 ug/l
 RT: 13.379 min Scan# 1885
 Delta R.T. -0.000 min
 Lab File: VW031853.D
 Acq: 14 Jul 2025 16:37




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 73535
 Ion Ratio Lower Upper
 105 100
 134 18.3 9.6 28.8

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
 Supervised By :Mahesh Dadoda 07/16/2025




#86
 p-Isopropyltoluene
 Concen: 49.837 ug/l
 RT: 13.494 min Scan# 1904
 Delta R.T. -0.000 min
 Lab File: VW031853.D
 Acq: 14 Jul 2025 16:37



Tgt Ion:119 Resp: 607316
 Ion Ratio Lower Upper
 119 100
 134 24.4 12.9 38.7
 91 24.4 12.7 38.0





#87

1,3-Dichlorobenzene

Concen: 48.669 ug/l

RT: 13.494 min Scan# 1904

Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 30244

Ion	Ratio	Lower	Upper
146	100		
111	43.5	22.3	66.9
148	62.8	31.8	95.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#88

1,4-Dichlorobenzene

Concen: 45.559 ug/l

RT: 13.580 min Scan# 1918




Delta R.T. 0.006 min


Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Tgt Ion:146 Resp: 289785


Ion	Ratio	Lower	Upper
146	100		
111	44.5	20.4	61.3
148	67.0	30.2	90.6





#89
n-Butylbenzene
Concen: 49.104 ug/l
RT: 13.818 min Scan# 1957
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 581390
Ion Ratio Lower Upper
91 100
92 54.0 27.2 81.5
134 24.1 12.2 36.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#90
Hexachloroethane
Concen: 47.250 ug/l
RT: 14.086 min Scan# 2001
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Tgt Ion:117 Resp: 103878
Ion Ratio Lower Upper
117 100
201 59.4 31.1 93.3





#91
1,2-Dichlorobenzene
Concen: 48.468 ug/l
RT: 13.866 min Scan# 1965
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37




Instrument :
MSVOA_W
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 27529
Ion Ratio Lower Upper
146 100
111 46.6 21.7 65.1
148 60.8 29.8 89.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 47.742 ug/l
RT: 14.482 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VW031853.D
Acq: 14 Jul 2025 16:37



Tgt Ion: 75 Resp: 28754
Ion Ratio Lower Upper
75 100
155 69.9 37.2 111.6
157 96.4 45.6 136.9





#93

1,2,4-Trichlorobenzene

Concen: 49.778 ug/l

RT: 15.128 min Scan# 2172

Delta R.T. -0.000 min

Lab File: VW031853.D


Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC



Tgt Ion:180 Resp: 17436

Ion Ratio Lower Upper

180 100

182 89.3 48.6 145.9




145 33.9 17.3 51.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025

Supervised By :Mahesh Dadoda 07/16/2025



#94

Hexachlorobutadiene

Concen: 44.218 ug/l

RT: 15.232 min Scan# 2189

Delta R.T. 0.006 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37




Tgt Ion:225 Resp: 74794


Ion Ratio Lower Upper

225 100

223 62.2 32.6 97.8

227 61.5 37.3 111.8





#95

Naphthalene

Concen: 53.647 ug/l

RT: 15.360 min Scan# 2210

Delta R.T. 0.006 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Instrument :

MSVOA_W

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 46071

Ion	Ratio	Lower	Upper
128	100		
127	13.0	10.8	16.2
129	11.0	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 07/15/2025
Supervised By :Mahesh Dadoda 07/16/2025



#96

1,2,3-Trichlorobenzene

Concen: 49.482 ug/l

RT: 15.549 min Scan# 2241

Delta R.T. -0.000 min

Lab File: VW031853.D

Acq: 14 Jul 2025 16:37

Tgt Ion:180 Resp: 159265

Ion	Ratio	Lower	Upper
180	100		
182	93.1	46.4	139.1
145	37.2	17.8	53.3

