

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071524\
 Data File : VW029569.D
 Acq On : 15 Jul 2024 12:23
 Operator : SY/MD
 Sample : VW0715SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK521

Quant Time: Jul 16 00:11:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jul 15 23:59:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	300424	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	262294	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	123986	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	91918	18.940	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.760%
7) Chloroethane-d5	2.893	69	69305	19.833	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.320%
11) 1,1-Dichloroethene-d2	4.021	65	46106	21.156	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	84.640%
21) 2-Butanone-d5	7.075	46	44279	30.394	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	60.780%
24) Chloroform-d	7.648	84	184985	21.299	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.200%
26) 1,2-Dichloroethane-d4	8.301	65	104267	20.468	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.880%
32) Benzene-d6	8.270	84	369005	23.484	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.920%
36) 1,2-Dichloropropane-d6	9.270	67	116677	23.224	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.880%
41) Toluene-d8	10.325	98	324793	23.504	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	10.575	79	40707	19.134	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.520%
47) 2-Hexanone-d5	10.922	63	29483	32.295	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.600%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	75065	18.403	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	73.600%
66) 1,2-Dichlorobenzene-d4	13.848	152	98457	22.802	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.200%
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	18812	3.498	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071524\
Data File : VW029569.D
Acq On : 15 Jul 2024 12:23
Operator : SY/MD
Sample : VW0715SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK521

Quant Time: Jul 16 00:11:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Jul 15 23:59:48 2024
Response via : Initial Calibration

