

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071618\
 Data File : VW003953.D
 Acq On : 16 Jul 2018 16:12
 Operator : VA/AP
 Sample : VW0716SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0716SBL01

Quant Time: Jul 16 16:41:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	253445	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	417830	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	383955	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	188637	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	154643	54.59	ug/l	0.00
Spiked Amount			Recovery	=	109.18%	
35) Dibromofluoromethane	7.88	113	131462	49.04	ug/l	0.00
Spiked Amount			Recovery	=	98.08%	
50) Toluene-d8	10.32	98	503018	48.36	ug/l	0.00
Spiked Amount			Recovery	=	96.72%	
62) 4-Bromofluorobenzene	12.62	95	170605	45.48	ug/l	0.00
Spiked Amount			Recovery	=	90.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.12	43	9128	13.41	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071618\
Data File : VW003953.D
Acq On : 16 Jul 2018 16:12
Operator : VA/AP
Sample : VW0716SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0716SBL01

Quant Time: Jul 16 16:41:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

